



National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
981-0300 / 981-0400 local 1157
<http://www.nkti.gov.ph/>

SUPPLEMENTAL BID BULLETIN NO. 20-107-1

**ONE (1) LOT SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF ONE (1)
UNIT HIGH-END AMBULANCE WITH CONVERSION REQUIREMENT**

Bid Reference ITB No. 20-107

This Supplemental Bid Bulletin **No. 20-107-1** is being issued to inform all the prospective bidders on the following amendment:

a. Section VII. Technical Specifications

Please see attached Revised Technical Specifications

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 17th day of November 2020 in Quezon City.

(SGD)JOSEPH MICHAEL A. JARO, MD
BAC, Chairman

Received by: (PLS SIGN) _____

Bidder's Name: (PLS PRINT) _____

Date: _____



Section VII.

Technical Specifications

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

Particulars	Requirements	Statement of Compliance
HIGH-END AMBULANCE: <i>The high-end ambulance should have at least enough width and height and in emergency situations would have enough space to allow emergency operations related to Donor Management. The minimum requirement of the dimension has to be complied, either by actual or true dimension of the car body or through modification of the car body by using a fiber glass material to suit the requirement.</i>		
External and Internal Dimension	At least 5,900mm x 2,000 x 2,500mm (L x W x H) with cabin height at least 72inches	
Transmission	At least 6-Speed Manual or Automatic	
Engine Technology	Direct Injection Four Stroke Common Rail Turbo Diesel or latest Model	
Engine Displacement	2.2L – 2.5L or better	
Fuel System	Direct Injection Common Rail Turbo Diesel or latest model	
Window	Power Window at Driver and Passenger	
Safety Features	At least the following safety features: dual front airbags for driver and passenger and standard safety harness, side impact beams , Anti-Lock braking system ABS, Rear parking assist and camera, Anti-theft alarm with immobilizer	
Audio System	As per manufacturer standard – radio mp3 with at least 4 speakers	
Door Requirement	Swing-out Double Rear Door with glass windows and Single Side-Sliding or Swing-out Door opposite to driver	
Standard Accessories	Manufacturer standard	



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
 East Avenue, Quezon City 1100
 8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Tires and wheels	At least R16 tubeless tires with alloy mags and spare	
ADDITIONAL REQUIREMENT FOR CONVERSION TO AMBULANCE		
Air-conditioning System	Dual Air-conditioning systems for driver and patient compartments	
Heavy Duty Power Inverter / Charger with deep cycle battery	Provide enough power supply for simultaneous tapping of equipment at least 2 universal sockets with at least 2,000watts inverter capacity and at least 100amp/hours battery with compatible 220volt battery charger at least 30 amps with fast and slow options and automatic functions	
Extra Light	Provide extra LED lighting at passenger area with switch accessible at same place, lighting must be operational even when the engine is turned off	
Semi-Automatic Rescue Stretcher	Max. load Capacity: 250 Kg., with adjustable side bars, stable frames, aluminum lightweight collapsible stretcher (stretcher inside the ambulance when at rest must be extendable to a certain height)	
First Aid Box	Polymer unbreakable PVC or Plastic, shock proof, weatherproof, water resistant	
Medicine/Equipment Cabinet	EQUIPMENT CABINET: Left side to front walls full size (height) cabinet made of Plywood marine or PVC board; cabinet divisions/partitions spacious enough to accommodate all equipment (dimension to be provided by NKTi) with complete fastening system	
Communication Equipment	Intra-vehicle intercom system (between driver and patient compartment)	
	DOH connected base radio transceiver (license and permits to be arranged by NKTi)	
	3-seater bench with 2-point seat belts and underneath cabinet	
	IV holder (Fixed on the head side of the patient)	
	Flooring: checkered stainless steel or aluminum flooring with washable non-slip rubber matting	
Emergency Lights and Siren	Red flashing LED lights on front top of the vehicle	
	Red flashing LED lights at upper back of the Ambulance	
	Red LED flashing lights on the front grill of the Ambulance	



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
 East Avenue, Quezon City 1100
 8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

	White strobe blinkers on all sides of the ambulance	
	Electric Siren with Four (4) Tones with standard medical ambulance tunes	
	Public Address (PA) System	
Oxygen Tank with Regulator and Tank	Provide 2-sets Size "F" Flask Type with fixed and safe slot with regulator, flow meter, face mask and other accessories; and provide additional 2-sets Size smaller than "F" Flask Type with holder attached to the stretcher complete with regulator, flow meter, face mask and other accessories. All medical oxygen tanks must be specific for ambulance use. The operation shall be manual switch from active oxygen tank to standby oxygen tank.	
	Additional Requirement: a) At least 20 pounds capacity each tank b) Must be color green c) Oxygen therapy set i. Gauge reading: 1,800 psi or higher ii. Flowmeter: 0 to 15 li/min iii. Humidifier bottle: Hard plastic, capacity: 200ml to 300ml iv. Accessories: mouthpiece, tubing	
Nebulizer	Diaphragm type pump assembly	
	Replaceable air filter	
	Compartment for storing the nebulizer kit and air tube	
	Power supply: 220V, 60Hz	
	Compressor pressure: at least 20 psig	
	Standard accessories: Two mouthpiece with tubing, two adult mask and two pedia mask, Air filters	
Scoop stretcher	Length: at least 200 cm	
	Adjustable length	
	Width: at least 42 cm	
	Loading capacity: at least 150 kg	
	The stretcher must have at least three (3) quick release patient restrain system	
	Frame material: high grade aluminum	



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
 East Avenue, Quezon City 1100
 8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Heavy Duty Stethoscopes (1 pediatric and 1 adult)	Two-sided Chest piece	
	Stainless Steel Chest Piece	
	Combination diaphragm and bell type	
	Rubberized ear tips	
	Standard Accessories: i. Spare Ear tip - 2 pcs. each for pediatric and adult ii. Spare Diaphragm and ring - 1 pc. Each for pediatric and adult	
Fire extinguisher	(rating 2A10BC)	
Heavy duty LED flashlight	with at least 1,000 lumens of brightness with extra rechargeable batteries (18650 at least 2000mah)	
Others	Kendrick Extrication Device	
	C-collar	
	Sam Splint	
	BP apparatus wall mounted	
	Ambubag for adult and pedia	
	Minor Surgical kit	
	I.V. hook with strap	
	2-seater with backrest foldable with seatbelt	
	3-seater with backrest foldable with seatbelt	
Body Requirement	Customized Body paint and Signage (verify with NKTi)	
Patient Compartment	At least 6ft. high (customized height)	
Statutory Requirement	Full Compliance with DOH Admin Order No. 2018-001	

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____