



**SUPPLEMENTAL BULLETIN NO. 21-012-1**

***Supply and Delivery of Two (2) units Defibrillator, Bi Phasic***

**Bid Reference BFO No. 21-012**

This Supplemental Bulletin No. 21-012-1 is being issued to clarify, modify and amend items/specifications in the Documents and to address key issues raised during the negotiation with suppliers for the afore-cited project, to wit:

**A. Changes in some items Technical Specifications**

Particulars	Amendment
<b>CONTENTS OF ELIGIBILITY AND TECHNICAL BLUE FOLDER</b>	<b>CONTENTS OF ELIGIBILITY AND TECHNICAL BLUE FOLDER</b>
Item No. 13  List of <b>COMPLETED</b> local brand installations in at least two (2) hospitals or health facilities excluding NKTi with contact person and phone numbers	Item No. 13  List of <b>COMPLETED local installations of the brand</b> in at least two (2) hospitals or health facilities excluding NKTi with contact person and phone numbers
<b>TECHNICAL SPECIFICATIONS</b>	<b>REVISED TECHNICAL SPECIFICATION</b>
For item:  Bi-Phasic waveform defibrillation with less than 5 seconds to charge	For item:  Bi-Phasic waveform defibrillation with less than <b>7</b> seconds to charge
	Please see attached Revised Technical Specification

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 6<sup>th</sup> day of August 2021 in Quezon City.

**VIOLETA M. VALDERRAMA, MD**  
**SBAC Chairman**



**National Kidney and Transplant Institute**  
**Special Bids and Awards Committee**  
 East Avenue, Quezon City 1100  
 981-0300 / 981-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## Section VII.

# Revised Technical Specifications

**Instruction:**

*Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.*

Item	Specification	Statement of Compliance
<b>1</b>	<b>Supply and Delivery of Two (2) units Defibrillator, Bi Phasic</b>	
	Must not exceed 8 kgs without battery	
	Bi-Phasic waveform defibrillation with less than 7 seconds to charge	
	With audible alarm on critical values for all parameters (high and low)	
	Defibrillator with Automated External Defibrillator (AED) MODE and manual MODE with voice prompt	
	With non-invasive pacing, ECG and arrhythmia monitoring, SPO2 pulse oximetry, 3 or 5 or 6 or 10-lead ECG	
	With non-invasive blood pressure monitoring, end-tidal CO2 for both intubated and non-intubated patients	
	With high resolution, Minimum of 5.5 inches TFT color LED/LCD Monitor	
	With external adult/pedia paddles, built-in 50mm ECG strip printer roll type, with SPO2 pulse oximetry with at least three (3)	

	reusable sensors (2 adult and 1 pedia) and compatible with disposable SPO2 sensor	
	With non-invasive pacing therapy capability	
	Electrical supply: 220 VAC $\pm$ 10%, Single phase with 60Hz OR Autovolt with GND	
	With built-in rechargeable lithium ion or nickel battery with at least four (4) hours capacity on full monitoring good for at least twenty (20) successive defibrillator maximum bi-phasic shocks	
	With medical grade or OEM trolley with basket for accessories, powder coated steel table stand with secure anchoring system clip or bolt, with at least four (4) Polyurethane caster wheels, with at least two (2) wheels with brake	
	<b>OTHER REQUIREMENTS:</b>	
	ECG Paper: at least five (5) 50mm rolls per unit	
	At least five (5) sets of disposable and compatible transcutaneous pads per unit	
	At least six (6) SPO2 adult disposable sensors per unit	
<p><b>Conforme:</b></p> <p>Name: _____</p> <p>Legal Capacity: _____</p> <p>Signature: _____</p> <p>Duly authorized to sign the Bid for and behalf of: _____</p>		