



National Kidney and Transplant Institute  
Special Bids and Awards Committee  
East Avenue, Quezon City 1100  
981-0300 / 981-0400 local 1157  
<http://www.nkti.gov.ph/>

## SUPPLEMENTAL BID BULLETIN NO. 21-036-1

### *Supply, Delivery, Testing and Commissioning of Two (2) units Refrigerated Centrifuge with Bundled Accessories*

#### Bid Reference IB No. 21-036

This Supplemental Bid Bulletin No. 21-036-1 is being issued to clarify, modify and amend items/specifications in the Bidding Documents for the aforementioned project, to wit:

#### A. Changes in some items Technical Specifications

Particulars	Changes
<b>Technical Specifications</b>	<b>Revised Technical Specification</b>
Touch screen for information display and control input	Touch screen <b>OR soft keys</b> for <b>full</b> information display and control input
Power supply: 100 - 240VAC or better, 60Hz	Power supply: <b>220 +/- 10% VAC, 60Hz or AUTOVOLT 50/60Hz</b>
One (1) unit appropriate Automatic Voltage Regulator (AVR)	One (1) unit appropriate Automatic Voltage Regulator (AVR)
One (1) unit digital portable balance	One (1) unit digital portable balance, <b>at least stainless-steel surface or ceramic surface: capacity at least 1kl</b>
With 4 pieces buckets with adaptors (4 pcs each) for 10ml, 15ml, 50ml tubes; blood bag with bio-containment lid	With 4 pieces buckets with adaptors (4 pcs each) for 10ml, 15ml, 50ml tubes; blood bag <b>holder with at least 300mL capacity</b> with bio-containment lid
	<b>Please see attached Revised Technical Specification</b>

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 18<sup>th</sup> day of January 2021 in Quezon City.

  
**JOSEPH MICHAEL A. JARO, MD**  
SBAC Chairman



Republic of the Philippines  
National Kidney and Transplant Institute  
Bids and Awards Committee  
East Avenue, Quezon City 1100  
8981-0300 / 9881-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## ***Section VII***

### ***Revised Technical Specifications***

***Instruction:***

*Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.*

<b>Item</b>	<b>Specification</b>	<b>Statement of Compliance</b>
<b>1</b>	<b><i>Supply, Delivery, Testing and Commissioning of Two (2) units Refrigerated Centrifuge with Bundled Accessories</i></b>	
	Table top model	
	Temperature Range: -10°C to +40°C	
	Microprocessor-controlled, programmable memory	
	With automatic imbalance detection	
	With visual and/or audible alarm	
	Digital display on the following:	
	- Time	
	- Overspeeding	

	- Temperature	
	- Door interlock	
	- System check/program error	
	Touch screen <b>OR soft keys</b> for <b>full</b> information display and control input	
	With stainless steel/aluminum swinging bucket rotor	
	Speed requirement: at least 3750 rpm for plate carrier; at least 4000 rpm for bucket	
	Refrigeration: CFC free	
	Maintenance-free brushless drive motor	
	Power supply: <b>220 +/- 10% VAC, 60Hz or AUTOVOLT 50/60Hz</b>	
	<b>BUNDLED INCLUSIONS/ACCESSORIES PER UNIT</b>	
	One (1) unit appropriate Automatic Voltage Regulator (AVR)	
	One (1) unit online UPS appropriate to the unit	
	One (1) unit digital portable balance, <b>at least stainless-steel surface or ceramic surface: capacity at least 1kl</b>	
	With 4 pieces buckets with adaptors (4 pcs each) for 10ml, 15ml, 50ml tubes; blood bag <b>holder with at least 300mL capacity</b> with bio-containment lid	
	With 4 pieces plate carriers, which can accommodate standard microplates, 4 HLA plates	

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_