



National Kidney and Transplant Institute
Special Bids and Awards Committee
East Avenue, Quezon City 1100
981-0300 / 981-0400 local 1157
<http://www.nkti.gov.ph/>

SUPPLEMENTAL BID BULLETIN NO. 21-064-2

***Supply, Delivery, Testing and Commissioning of One (1) unit Rotary
Microtome with Accessories***

Bid Reference IB No. 21-064

This Supplemental Bid Bulletin No. 21-064-1 is being issued for the correction of the schedule of Pre-Bid Conference as posted in the PhilGEPS for the aforementioned project, to wit:

Particulars	Amendment
Bid Data Sheet (Clause 21.2)	Revised Bid Data Sheet (Clause 21.2)
Item No. 4 Certificate for at least Five (5) years warranty on parts, unit and service including annual calibration and preventive maintenance.	Item No. 4 Certificate for at least Five (5) years warranty on parts, unit and service including third party items with annual calibration and preventive maintenance or manufacturer specifications
Item No. 5 Certificate of availability of spare parts and accessories for five (5) years after unit has ceased production.	Item No. 5 Certificate of availability of spare parts and accessories for five (5) years after the warranty period
Item No. 9 List of COMPLETED similar brand local installations by the winning bidder in at least two (2) hospitals or health or research facilities excluding NKTi with contact person and phone numbers.	Item No. 6 (Additional) To submit list of consumables and accessories with price schedule
Item No. 10 Certificate of Satisfactory Performance from at least three (3) clients issued and signed by the Authorized Officer of the Laboratory Department or Section or the Head of Agency including NKTi with letterhead and contact number	Item No. 10 List of COMPLETED similar brand and similar functionality local installations by the winning bidder in at least two (2) hospitals or health or research facilities excluding NKTi with contact person and phone numbers
	Item No. 11 Certificate of Satisfactory Equipment Performance within the last three (3) years from at least three (3) local or domestic clients issued and signed by the Authorized Officer of the Laboratory Department or Section or the Head of Agency including NKTi, if applicable with letterhead and contact number

<p>Item No. 12 Certificate of ISO 13485 or its international equivalent for healthcare equipment for the offered equipment</p> <p>Item No. 13 Certificate of guarantee that the bidder shall provide a demo unit (same unit/model offered in the bid) within a non-extendable five (5) calendar days upon receipt of notification or a site visit within NCR or live virtual demonstration of the unit offered under recording.</p>	<p>Item No. 13 Certificate of ISO 13485 and IEC/EC/EEC or its international equivalent for healthcare equipment for the offered equipment</p> <p>Item No. 14 Certificate of guarantee that the bidder shall provide, in the order of priority, one of the following within a non-extendable five (5) calendar days upon receipt of notification:</p> <ol style="list-style-type: none"> 1. a demo unit (same unit/model offered in the bid) or 2. arrange a site visit within NCR or 3. live virtual demonstration of the unit offered under recording
<p>Technical Specifications</p>	<p>Revised Technical Specifications</p>
<p>Floataion bath (same brand)</p>	<p>Floataion bath Type: Bench-Top paraffin floataion bath Dimensions: (maximum requirement): L 1.3in x W 16in LED Display panel With slide warmer on at least one side Operating Temperature Range: +15°C to +60°C or better IP Protection Class (IEC 60529): IP20 or better Power Requirement: 220 +/- 10% VAC, 60Hz or AUTOVOLT</p> <p>Please see attached Revised Technical Specifications</p>

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 22nd day of March 2021 in Quezon City.

JOSEPH MICHAEL A. JARO, MD
SBAC Chairman



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VII.

Revised Technical Specifications

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

Item	Specification	Statement of Compliance
	Type: Table-Top Mechanical or Semi-Automated If semi-automated, must have mechanical adjustment	
	Width: Maximum of 480mm	
	Depth: Maximum of 650mm	
	Height (without top tray): Maximum of 350mm	
	Cutting thickness adjustments: 1 – 60um or better 1um increments (or better)	
	Trimming section thickness: 10 – 30um coarse feed or better	
	Vertical Stroke: At least 70mm	
	Horizontal Travel: 24mm or better	

	Specimen Retraction: With automatic retraction	
	Specimen clamp and orientation: Quick release with micro adjustment orientation at least 8° (Horizontal and Vertical) and 360 (Z-axis) adjustments	
	Maximum specimen size: 55mm x 50mm x 30mm	
	Handwheel Lock/Brake: With mechanical lock or better	
	Blade Holder: For low profile blade. Blade holder can be moved sideways both to the left and right	
	Accessories:	
	Floatation bath (same brand) Type: Bench-Top paraffin floatation bath Dimensions: (maximum requirement): L 1.3in x W 16in LED Display panel With slide warmer on at least one side Operating Temperature Range: +15°C to +60°C or better IP Protection Class (IEC 60529): IP20 or better Power Requirement: 220 +/- 10% VAC, 60Hz or AUTOVOLT	
	With waste tray	
	Brush for Waste	
	300 pcs low profile blade	
	1 unit gooseneck LED table lamp	

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____