



National Kidney and Transplant Institute
Special Bids and Awards Committee
East Avenue, Quezon City 1100
981-0300 / 981-0400 local 1157
<http://www.nkti.gov.ph/>

SUPPLEMENTAL BID BULLETIN NO. 21-064-3

***Supply, Delivery, Testing and Commissioning of One (1) unit Rotary
Microtome with Accessories***

Bid Reference IB No. 21-064

This Supplemental Bid Bulletin No. 21-064-3 is being issued to clarify, modify and amend items/specifications in the Bidding Documents for the aforementioned project, to wit:

Particulars	Amendment
Technical Specifications	Revised Technical Specifications
Floatation bath Type: Bench-Top paraffin floatation bath Dimensions: (maximum requirement): L 1.3in x W 16in LED Display panel With slide warmer on at least one side Operating Temperature Range: +15°C to +60°C or better IP Protection Class (IEC 60529): IP20 or better Power Requirement: 220 +/- 10% VAC, 60Hz or AUTOVOLT	Floatation bath Type: Bench-top paraffin floatation bath Dimensions (maximum requirement): L 15in x W 16in LED display panel Digital temperature controlled With slide warmer on at least one side Operating temperature range: +15°C to +6°C or better IP protection class (IEC 60529): IP20 or better Power requirement: 220 +/- 10% VAC, 60Hz or AUTOVOLT Please see attached Revised Technical Specifications

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 26th day of March 2021 in Quezon City.

(sgd.) JOSEPH MICHAEL A. JARO, MD
SBAC Chairman



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VII.

Revised Technical Specifications

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

Item	Specification	Statement of Compliance
	Type: Table-Top Mechanical or Semi-Automated If semi-automated, must have mechanical adjustment	
	Width: Maximum of 480mm	
	Depth: Maximum of 650mm	
	Height (without top tray): Maximum of 350mm	
	Cutting thickness adjustments: 1 – 60um or better 1um increments (or better)	
	Trimming section thickness: 10 – 30um coarse feed or better	
	Vertical Stroke: At least 70mm	
	Horizontal Travel: 24mm or better	

	Specimen Retraction: With automatic retraction	
	Specimen clamp and orientation: Quick release with micro adjustment orientation at least 8° (Horizontal and Vertical) and 360 (Z-axis) adjustments	
	Maximum specimen size: 55mm x 50mm x 30mm	
	Handwheel Lock/Brake: With mechanical lock or better	
	Blade Holder: For low profile blade. Blade holder can be moved sideways both to the left and right	
	Accessories:	
	Floatation bath Type: Bench-top paraffin flotation bath Dimensions (maximum requirement): L 15in x W 16in LED display panel Digital temperature controlled With slide warmer on at least one side Operating temperature range: +15°C to +6°C or better IP protection class (IEC 60529): IP20 or better Power requirement: 220 +/- 10% VAC, 60Hz or AUTOVOLT	
	With waste tray	
	Brush for Waste	
	300 pcs low profile blade	
	1 unit gooseneck LED table lamp	

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____