



National Kidney and Transplant Institute
Special Bids and Awards Committee
East Avenue, Quezon City 1100
981-0300 / 981-0400 local 1157
<http://www.nkti.gov.ph/>

SUPPLEMENTAL BID BULLETIN NO. 21-107-1

***Re-bidding for the Supply, Delivery, Testing and Commissioning of One (1)
unit Gross Station with Accessories***

Bid Reference IB No. 21-107

This Supplemental Bid Bulletin No. **21-107-1** is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

A. Changes in some items Technical Specifications

Particulars	Amendment
Bid Data Sheet (Clause 21.2)	Revised Bid Data Sheet (Clause 21.2)
Item No. 9 Certificate of guarantee issued by the bidder that in case of breakdown and on-going repair in excess of 24 hours, the bidder shall provide a service unit with the same functionality.	Item No. 9 DELETED
Technical Specifications	Revised Technical Specification
Item: Dimension: L 60-72in x W 24-32in x H 70-90in	Item: Dimension: L 60-73in x W 24-32in x H not more than 96in or equivalent
Item: Cutting boards: L40 to 60cm x W30 to 50cm x 1.5 to 2cm thickness Polyethylene dissecting board any color but preferably blue included in the gross station.	Item: Cutting boards: Medical Grade Manufacturer's Specifications Please see attached Revised Technical Specification

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 13th day of July 2021 in Quezon City.

VIOLETA M. VALDERRAMA, MD
SBAC Chairman



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VII.

Revised Technical Specifications

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

Item	Specification	Statement of Compliance
	Type: Floor Type	
	Dimension: L 60-73in x W 24-32in x H not more than 96in or equivalent	
	Adjustable Height: Electronic or Hydraulic	
	Material: Stainless steel, Medical Grade 304. satin or hairline finished	
	Ventilation: Appropriate Replaceable Filters for gross station, including ventilation system for the cutting area, 6-8 ACH, ducting will be from ceiling to the upper floor	
	Faucet control: Infrared sensor or foot pedal control	
	Effective Top Working Area (L x D): 58-72in x 17-32in (basin is part of the working area)	
	With Basin positioned on the right or left.	

	<p>Integrated PC workstation:</p> <ul style="list-style-type: none"> At least 6 cores latest generation processor At least 500GB SSD for operating system At least 1 TB HDD internal storage 32GB RAM At least 6GB 256bits GPU 24" 1920x1080 IPS 144Hz Adjustable monitor Latest Professional OS Appropriate Online UPS 	
	<p>CPU, monitor and keyboard mounting:</p> <ul style="list-style-type: none"> Articulating/ fix CPU, monitor and keyboard mount 	
	<p>Connectivity: Network ready via LAN and WIFI</p>	
	<p>Alarm: Audible Alarm</p>	
	<p>Cutting boards: Medical Grade Manufacturer's Specifications</p>	
	<p>Flex lights: LED flex lights</p>	
	<p>Imaging system: Dedicated or built-in camera with high resolution of at least 16 mega pixel with imaging software (capable of measurement and image editing)</p>	
	<p>Other features</p> <ul style="list-style-type: none"> Magnetic Instrument strip Tissue dye holder Voice recorder at least 64GB capacity Microphone on flex arm 	
	<p>Electronic scale platform: Bench scale, 3kg maximum capacity, 0.1g certified readability</p>	
	<p>With rinsing mechanism</p>	
	<p>Filter: Activated carbon filter</p>	
	<p>Electrical requirement: 220 +/- 10% VAC, 60Hz or AUTOVOLT</p>	

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____