



## SUPPLEMENTAL BID BULLETIN NO. 21-108-1

### *Supply, Delivery, Testing and Commissioning of Two (2) units Video Laryngoscope*

#### Bid Reference IB No. 21-108

This Supplemental Bid Bulletin No. 21-108-1 is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

**A. Changes in some items Technical Specifications**

Particulars	Amendment
<b>Bid Data Sheet (Clause 21.2)</b>	<b>Revised Bid Data Sheet (Clause 21.2)</b>
Item No. 12 Certificate of guarantee issued by the bidder that in case of breakdown and on-going repair, the bidder shall provide a service unit with a similar unit	Item No. 12 Certificate of guarantee issued by the bidder that in case of breakdown and on-going repair, the bidder shall provide a service unit with a similar <b>functionality</b>
Item No. 14 Certificate of guarantee that the bidder shall provide, in the order of priority, one of the following within a non-extendable five (5) calendar days upon receipt of notification: 1. a demo unit (same unit/model offered in the bid); or 2. arrange a site visit within NCR; or 3. live virtual demonstration of the unit offered under recording	Item No. 14 Certificate of guarantee that the bidder shall provide demonstration of the same unit offered in the bid within a non-extendable five (5) calendar days upon receipt of notification
<b>Schedule of Requirements</b>	<b>Revised Schedule of Requirements</b>
Delivered, Weeks/Months: One Hundred Sixty (60) calendar days	Delivered, Weeks/Months: Sixty (60) calendar days.  <b>Please see attached Revised Schedule of Requirements</b>
<b>Technical Specifications</b>	<b>Revised Technical Specification</b>
<b>ACCESSORIES OF MONITOR</b>	<b>ACCESSORIES OF MONITOR (for each unit)</b>
with crossbar for mounting of monitor to the stand	with crossbar for mounting of monitor to the stand
Medical Grade Stand with Minimum of 4 caster wheel with two breaks	Medical Grade Stand with Minimum of 4 caster wheel with two breaks
With stand -at least 120cm height	With stand -at least 120cm height
With stainless steel holder/tray for blades and module	With stainless steel holder/tray for blades and module  <b>With blade stand holder</b>
<b>MICINTOSH BLADES</b>	<b>MACINTOSH BLADES (for each unit)</b>
Size: 3	Size: 3
Blade Tip: Always under direct view for safe navigation	Blade Tip: Always under direct view for safe navigation
Button: For image capturing and recording with indicator light	Button: For image capturing and recording with indicator light
Protection Cap: With protection cap for sterilization	Protection Cap: With protection cap for sterilization
	<b>OTHER ACCESSORY (for each unit)</b>
	<b>Defogger or anti-fog</b>
	<b>Please see attached Revised Technical Specifications</b>

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 16<sup>th</sup> day of July 2021 in Quezon City.

**VIOLETA M. VALDERRAMA, MD**  
**SBAC Chairman**



Republic of the Philippines  
National Kidney and Transplant Institute  
Bids and Awards Committee  
East Avenue, Quezon City 1100  
8981-0300 / 9881-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## ***Section VII.***

### ***Revised Technical Specifications***

***Instruction:***

*Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.*

<b>Item</b>	<b>Specification</b>	<b>Statement of Compliance</b>
	<b>MONITOR</b>	
	With data storage SD card At least 8gb	
	With USB access	
	Picture and video recording and real time capturing	
	HDMI Video output	
	Screen size At least 7"	
	with playback of recorded video clips and still images	
	Splash proof	
	Dual endoscope connection or inputs	
	Rechargeable battery	

	Shock resistant	
	Interoperable with existing video laryngoscope and Flexible Intubating Video Endoscope	
	<b>ACCESSORIES OF MONITOR (for each unit)</b>	
	With crossbar for mounting of monitor to the stand	
	Medical Grade Stand with Minimum of 4 caster wheel with two breaks	
	With stand -at least 120cm height	
	With stainless steel holder/tray for blades and module	
	With blade stand holder	
	<b>MACINTOSH BLADES (for each unit)</b>	
	Size: 3	
	Blade Tip: Always under direct view for safe navigation	
	Angle of View: At least 45 degrees	
	Blade Tip: Reusable	
	Button: For image capturing and recording with indicator light	
	Protection Cap: With protection cap for sterilization	
	<b>POWER SUPPLY</b>	
	110-240 V, 50-60Hz	
	<b>OTHER ACCESSORY (for each unit)</b>	
	Defogger or anti-fog	

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_



Republic of the Philippines  
National Kidney and Transplant Institute  
Bids and Awards Committee  
East Avenue, Quezon City 1100  
8981-0300 / 9881-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## Section VI.

### *Revised Schedule of Requirements*

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Qty./ UOM	Total	Delivered, Weeks/Months	Place of Delivery
1	Supply, Delivery Testing and Commissioning of Video Laryngoscope	2 units		Sixty (60) calendar days	NKTI SMD

**Standard Payment Terms:** Payments shall be made promptly by NKTI, within sixty (60) days from submission of complete documents *i.e.* invoice or claim by the Supplier.

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_