



SUPPLEMENTAL BID BULLETIN NO. 21-109-1

**Bidding for the
Supply and Delivery of Sixty-Eight Thousand One Hundred Eighty-Two
Pieces (68,182 pcs.) Particulate Filter Mask, Face**

Bid Reference IB No. 21-109

This Supplemental Bid Bulletin No. 20-109-1 is being issued to clarify, modify and amend items/specifications in the Bid Documents in response to the request for clarification from prospective bidders and to confirm key issues addressed during the Pre-Bid Conference on July 28, 2021 for the aforementioned project.

a. Change in the Title of the Project:

Original	Amended:
Supply and Delivery of Sixty Thousand Pieces (60,000 pcs.) Particulate Filter Mask, Face	Supply and Delivery of Sixty-Eight Thousand One Hundred Eighty-two Pieces (68,182 pcs.) Particulate Filter Mask, Face

b. **Section VII. TECHNICAL SPECIFICATIONS** is hereby amended, attached as *Revised Technical Specifications*.

c. **Bid Proposals Form** is revised to reflect the **changes in the technical specifications**:

Item No.	Particulars	Specification
1.	Particulate Filter Mask, Face	Packaging: <ul style="list-style-type: none">- Usage: Single use, Disposable- Label: NIOSH name, TC approval number, Filter designation and model- Presentation: can be? individually packed or maximum of 30 pieces per pack. for distribution purposes. Specifications: <ul style="list-style-type: none">- Material: Latex free- Size: Standard/ regular/medium or its equivalent. Features: <ul style="list-style-type: none">- may be soft inner shell or flat fold, with no exhalation valve- with adjustable nose clip- with nose foam- straps or elastic bands should fit around the neck and back of the head- passed as a medical mask or fluid-resistant- provides at least 95% filtration to airborne particles, not resistant to oil Functionality: <ul style="list-style-type: none">- provided good seal in the face, with no air leakage- passed the recommended fit testing rate >80% (0.2 fit factor), to provided samples necessary for the fit testing (at least 25 pcs.)- with approved CDC or EUA and Philippine FDA.

d. Approved Budget Cost is hereby amended:

ORIGINAL:				
Particular	UOM	TOTAL QTY.	NKTI APPROVED BUDGET COST WITH VAT	
			UNIT COST	TOTAL COST
Particulate Filter Mask, Face	pcs.	60,000	132.32	7,939,200.00
AMENDED:				
VAT EXEMPTION UNDER BIR REVENUE MEMORANDUM CIRCULAR No. 81-2081			NKTI APPROVED BUDGET COST WITHOUT VAT	
Particular	UOM	TOTAL QTY.	UNIT COST	TOTAL COST
Particulate Filter Mask, Face	pcs.	68,182	116.44	7,939,112.08

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 4th day of August 2021 in Quezon City.

(sgd.) ARNOLD JOSEPH M. FERNANDEZ, MD
BAC, Chairman

Received by: (PLS SIGN) _____

Bidder's Name: (PLS PRINT) _____

Date: _____

Section VII.

Revised Technical Specifications

NKTI Reference No. IB 21-109

Revised Technical Specifications

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

Item No.	Particulars	Specification	Statement of Compliance
1	Particulate Filter Mask, Face	<p>Packaging:</p> <ul style="list-style-type: none"> - Usage: Single use, Disposable - Label: NIOSH name, TC approval number, Filter designation and model - Presentation: individually packed or maximum of 30 pieces per pack. <p>Specifications:</p> <ul style="list-style-type: none"> - Material: Latex free - Size: Standard/regular/medium or its equivalent <p>Features:</p> <ul style="list-style-type: none"> - may be soft inner shell or flat fold, with no exhalation valve - with adjustable nose clip - with nose foam - straps or elastic bands should fit around the neck and back of the head - passed as a medical mask or fluid-resistant - provides at least 95% filtration to airborne particles, not resistant to oil <p>Functionality:</p> <ul style="list-style-type: none"> - provided good seal in the face, with no air leakage - passed the recommended fit testing rate >80% (0.2 fit factor), to provided samples necessary for the fit testing (at least 25 pcs.) - with approved CDC or EUA and Philippine FDA. 	

We guarantee that we will submit sample within 3 calendar days upon notice from the BAC. Failure to submit within the prescribed period will automatically disqualify the bidder.

(NOTE: Submission of samples are no longer necessary if used by the end-user or otherwise specified by the end-user.)

(Upon receipt of notice, Lowest Calculated Bidder shall also submit clear copies of the colored picture and COMPLETE PRODUCT DATA SHEET [i.e., brochures, pamphlets, sales literature, documentation of standards of compliance, etc.], with the required specifications already highlighted, of the item proposed in A4 size photo paper indicating the brand name with reference number).

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

REVISED PRICE SCHEDULE

ANNEX A

IB NO. 21-109: Supply and Delivery of Sixty-Eight Thousand One Hundred Eighty-Two Pieces (68,182 pcs.) Particulate Filter Mask, Face

August 11, 2021, (Wednesday) 9:00AM

BAC Conference Room, G/F NKT Main Building

ITEM NO.	ITEM / SUPPLIES	TECHNICAL SPECIFICATION	UOM	TOTAL QTY.	NKT APPROVED BUDGET COST		SUPPLIER	BIDDER'S PROPOSAL		BRAND	REMARKS
					UNIT COST	TOTAL COST		UNIT COST	TOTAL COST		
1	Particulate Filter Mask, Face	<p>Packaging:</p> <ul style="list-style-type: none"> - Usage: Single use, Disposable - Label: NIOSH name, TC approval number, Filter designation and model - Presentation: individually packed or maximum of 30 pieces per pack. <p>Specifications:</p> <ul style="list-style-type: none"> - Material: Latex free - Size: Standard/ regular/medium or its equivalent. <p>Features:</p> <ul style="list-style-type: none"> - may be soft inner shell or flat fold, with no exhalation valve - with adjustable nose clip - with nose foam - straps or elastic bands should fit around the neck and back of the head - passed as a medical mask or fluid-resistant - provides at least 95% filtration to airborne particles, not resistant to oil <p>Functionality:</p> <ul style="list-style-type: none"> - provided good seal in the face, with no air leakage - passed the recommended fit testing rate >80% (0.2 fit factor), to provided samples necessary for the fit testing (at least 25 pcs.) - with approved CDC or EUA and Philippine FDA. 	piece	68,182	116.44	7,939,112.08					
TOTAL							7,939,112.08				

Printed Name of Company

Date

Signature

Address

Tel. No.

Printed Name and Designation