



## SUPPLEMENTAL BID BULLETIN NO. 21-111-1

### *Supply, Delivery, Testing and Commissioning of Various Nursing Medical Equipment (Line Bidding)*

#### Bid Reference IB No. 21-111

This Supplemental Bid Bulletin No. **21-111-1** is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

#### A. Changes in some items Technical Specifications

Particulars	Amendment
<b>Technical Specifications</b>	<b>Revised Technical Specification</b>
<b>HOSPITAL BED, ELECTRIC</b>	<b>HOSPITAL BED, ELECTRIC</b>
Weight Capacity : 22 kgs - 275kgs	<b>Weighing Scale Capacity: 275kgs or higher</b>
Bed Dimension : Overall Width : 30-45inches Overall Length : 80 -95inches	Bed Dimension : Overall Width : 30-45inches Overall Length : <b>78</b> -95inches
With detachable rack / holder back -up battery at least 150 cycles	With <b>Built-in</b> battery at least 150 cycles
Angle indicator display with weighing scale in kilograms (minimum if 20 kgs ) cardio pulmonary resuscitation (CPR ) release function	Angle indicator display with weighing scale in kilograms cardio pulmonary resuscitation (CPR ) release function
Base : over molded form polyurethane for mural protection / powder coated steel / high pressure laminated (HPL) mounted on (4) swiveling casters (diameter 50 - 80mm )	DELETED
Body : laminated chipboard panel / high pressure laminated with aluminum protection radius corner equipped on upper part with 2-3 drawers	DELETED
Drawers : laminated chipboard panel with rounded ABS edges and with stainless metal handles	DELETED
Top : Laminated with concealed hinges	DELETED
Bed Dimension : Overall Width : 30-45inches Overall Length : 80 -95inches	Dimension : <b>Depth 40cm - 47cm x Width 40cm-47cm x Height at least 80 cm</b>
Column, foot and board support plate in steel protected metallic epoxy coating	DELETED
<b>HOSPITAL BED, MANUAL</b>	<b>HOSPITAL BED, MANUAL</b>
Height: Low Position: Floor to top of deck 15-20"	Height: Low Position: Floor to top of deck <b>37cm</b> - 53cm or its equivalent

Equipped with four (4) at least 5" diameter polyurethane swivel wheel caster, two brakes (one on foot part and one on head part)	Equipped with four (4) swivel wheel with at least 5" diameter polyurethane caster, minimum of two brakes (one on foot part and one on head part)
Back section lifting angle: 0-80°	Back section lifting angle: up to 75°
Leg section lifting angle: 0-40°	Leg/Knee raise section lifting angle: 0-40°
Low Position lift to high position: 48cm-72cm	DELETED
Mattress	Mattress
At least 5" thickness	At least 4" thickness
With replaceable Leatherette Cover	With replaceable, dark colored Leatherette Cover
At least 90cm x 190cm Length	At least Width: 85cm x Length: 190cm must be appropriate with the offered external /overall dimensions of the bedframe
<b>Bedside Table:</b>	<b>Bedside Table: OEM</b>
Standard Size: Length/width: 42-50cm	Standard Size: Length/width: 42-50cm
Depth: 44-50cm	Depth: 44-50cm
Height: 75-92cm	Height: 75-92cm
<b>Overbed Table:</b>	<b>Overbed Table: OEM</b>
Adjustable (Up and down)	Adjustable via gas spring lever (Up and down)
<b>CARDIAC MONITOR WITH DEFIBRILLATOR AND EXTERNAL PACER</b>	<b>CARDIAC MONITOR WITH DEFIBRILLATOR AND EXTERNAL PACER</b>
Lightweight 5-7.5 kg	Lightweight 5-7.6 kg
with secure handle	with handle for portable transport
ECG: capable of at least 3 to 6 leads; with non-invasive pacing, ECG and arrhythmia monitoring and accessories	ECG: capable of 3 leads; with non-invasive pacing, ECG and arrhythmia monitoring and accessories
Pulse Oximeter - SP02 monitor 1 large (adult) and 1 pediatric reusable sensor (soft non adhesive, rubberized)	Pulse Oximeter - SP02 monitor 1 large (adult) and 1 pediatric reusable sensor (soft non adhesive)
with end tidal CO2	with end tidal CO2, either sidestream, microstream or mainstream
<b>Interoperability with existing defibrillator / monitor at the ICU / ER and all existing units at the ward</b>	DELETED
With stable stand and secure hook & loop fastener	With OEM stable stand and secure hook & loop fastener
Mobile, minimum of 5 pcs plastic wheel caster with at least one (1) wheel with brake	Mobile, minimum of four (4) pcs plastic wheel caster with at least one (1) wheel with brake

<b>VACUUM SUCTION, WALL REGULATOR WITH BOTTLE</b>	<b>VACUUM SUCTION, WALL REGULATOR WITH BOTTLE</b>
1600 to 2000 liters capacity plastic collection bottle	1600ml to 2000ml capacity plastic collection bottle
Autoclavable	Autoclavable bottle
With fitted tubing connector, at least 100cm in length	With fitted tubing connector, at least 40cm in length
With bacterial filter	With bacterial filter at least five (5) pieces per set
Regulator Range: 0 to 750mm Hg Easy to manipulate	Regulator Range: 0 to 760 mm Hg Easy to manipulate knob and lever control
With secure holder to be attached on the wall	With secure holder to be attached on the wall that will fit to existing wall suction lines
	<b>Please see attached Revised Technical Specifications</b>

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 9<sup>th</sup> day of August 2021 in Quezon City.

**VIOLETA M. VALDERRAMA, MD**  
**SBAC Chairman**



**Republic of the Philippines**  
**National Kidney and Transplant Institute**  
**Bids and Awards Committee**  
 East Avenue, Quezon City 1100  
 8981-0300 / 9881-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## Section VII.

### Revised Technical Specifications

**Instruction:**

*Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.*

Item	Specification	Statement of Compliance
1	<b>HOSPITAL BED, ELECTRIC</b>	
	Bed Frame: made of high quality type of powder coated steel	
	Bed Panel: made of high quality type of powder coated steel / polymer plastic	
	Head and Foot Board: removable and detachable ; light weight material ; made of high quality plastic such as polymer or acrylonitrile Butadiene Styrene (ABS )	
	Side Rails : Controls : a. Integrated in split b. Tuck-away side rails c. Material : made up of high quality plastic materials such as polymer or acrylonitrile butadiene styrene (ABS)	
	IV Pole Holders : with at least four (4) holders / hook with stainless steel/ medical grade stainless steel or plastic on each side of the bed	
	IV Pole Socket : Depth : 2-5 inches ; four corners of the bed	
	Drainage Bag Holder : both sides of the bed ; adjustable ; stainless steel / medical grade	

	Weighing Scale Capacity: 275kgs or higher	
	Bed Articulations / Control : Max head of bed elevation : 60 - 75 degrees	
	Max Knee Elevation : at least 20 - 45 degrees maximum Trendelenburg and reverse trend : at least 10-25 degrees	
	Bed Dimension : Overall Width : 30-45inches Overall Length : 78 -95inches	
	Four (4) corner plastic (Polyurethane) roller bumpers four (4) polyurethane caster locking wheels and with central brake with at least two (2) locking casters (back and central brake )	
	Power Supply : 220 VAC $\pm$ 10% 60Hz or Autovolt with appropriate AVR	
	With Built-in battery at least 150 cycles	
	Angle indicator display with weighing scale in kilograms cardio pulmonary resuscitation (CPR ) release function	
	With anti-pressure foam mattress dimension : 198 – 208cm x at least 85 x 14cm	
	With stainless nameplate bearing the name and contact number of the supplier	
	Bedside Cabinet : same brand with bed, OEM with 2 – 3 drawers	
	Dimension : Depth 40cm - 47cm x Width 40cm-47cm x Height at least 80 cm	
	Overbed Table : same brand with the bed, OEM	
	Height Adjustable from 73 to 108 cm	
	30 kg safe working load on top center and 16kg stability in every point	
	Foot Stool : single step ; made of stainless steel , non-slip rubber matting on top , tubular legs with rubber tip on each leg	
	IV Pole : stainless ; with 4 pivot hooks , with 5 roller wheel caster , height adjustable up to 250 cm	
2	<b>HOSPITAL BED, MANUAL</b>	
	Length: (bed frame only): 210cm – 230cm	
	Width: (bed frame only): 90cm – 100cm	
	Height: Low Position: Floor to top of deck 37cm - 53cm or its equivalent	
	High Position: Floor to top deck 65cm – 90cm or its equivalent	

	<b>Other Specifications:</b>	
	Corrosion resistant steel frame	
	Aluminum Folding Side Rails	
	Heavy Duty light weight polymer material or Acrylonitrile butadiene styrene (ABS) plastic Head and Foot Board, detachable	
	Steel powder coated bed surface with ventilate holes	
	Equipped with four (4) swivel wheel with at least 5" diameter polyurethane caster, minimum of two brakes (one on foot part and one on head part)	
	Carrying capacity: at least 150 - 240kg	
	Back section lifting angle: up to 75°	
	Leg/Knee raise section lifting angle: 0-40°	
	<b>Accessories:</b>	
	<b>IV Pole Holders:</b>	
	Four (4) holes: One (1) in each corner of the bed	
	With Two (2) detachable stainless IV pole with four (4) full U-shaped or spiral hooks that securely holds intravenous infusion bags or bottles	
	Adjustable height from at least 45-76cm	
	<b>Mattress:</b>	
	At least 4" thickness	
	With replaceable, dark colored Leatherette Cover	
	At least Width: 85cm x Length: 190cm must be appropriate with the offered external /overall dimensions of the bedframe	
	<b>Bedside Table: OEM</b>	
	Acrylonitrile butadiene styrene (ABS) plastic	
	With one (1) drawer	
	With one (1) cabinet	
	With four (4) Caster ball wheel with 2 brakes	

	Standard Size: Length/width: 42-50cm	
	Depth: 44-50cm	
	Height: 75-92cm	
	<b>Overbed Table: OEM</b>	
	Adjustable via gas spring lever (Up and down)	
	With four (4) caster ball wheel with 2 brakes	
	Made of high-quality plastic material such as polymer or Acrylonitrile Butadiene Styrene (ABS)	
	<b>Foot Stool:</b>	
	Stainless, with rubber stopper on each leg	
	Standard Size : at least 420mm x 280 x 215 (L x W x H)	
	Non-skid rubber matting on top	
3	<b>CARDIAC MONITOR WITH DEFIBRILLATOR AND EXTERNAL PACER</b>	
	Lightweight 5-7.6 kg	
	with handle for portable transport	
	Biphasic waveform	
	with audible alarm on critical values for all parameters (high and low)	
	defibrillator with automated external defibrillator (AED) Mode and manual mode	
	ECG: capable of 3 leads; with non-invasive pacing, ECG and arrhythmia monitoring and accessories	
	Pulse Oximeter - SP02 monitor 1 large (adult) and 1 pedia reusable sensor (soft non adhesive)	
	Non-invasive BP monitoring	
	with end tidal CO2, either sidestream, microstream or mainstream	
	High Resolution TFT Color LCD Monitor minimum of 6 inches	

	External Paddle : Adult and Pedia	
	Built in 50mm ECG Strip printer	
	internal data storage of at least 6 hours of continuous ECG waveforms and events	
	Power Supply : 220 VAC $\pm$ 10% 60Hz or Autovolt	
	With rechargeable battery with capacity gauge minimum of at least 5 hours continuous operation	
	With OEM stable stand and secure hook & loop fastener	
	Mobile, minimum of four (4) pcs plastic wheel caster with at least one (1) wheel with brake	
4	<b>ELECTRIC PILL CRUSHER</b>	
	Dimension: 240mm (L) x 164mm(D) x 253mm(H)	
	Weight: 5.11 kg	
	Battery Type: Ni MH Battery, 12V OC 3600mAH; rechargeable	
	Battery run time approximately: 2.75 hrs.	
	Battery charge time: approximately 4 hours	
	Pulverizes medications to a fine powder suitable for tube feeding	
5	<b>MEDICAL HOSPITAL INFUSION TROLLEY CART WITH DRAWERS, CENTRAL LOCK AND EXTENSION TABLE</b>	
	Designed to supply the medication when nurse make rounds to visit patient	
	Installed multi - bin containers at rear side to make easy and multipurpose storage	
	Top Table: polypropylene / stainless steel	
	Body Frame: aluminum / ABS plastic with 1-2 waste bin	
	Guard Rail: Aluminum	
	Drawer: Aluminum / ABS Plastic	



	Caster: four (4) bearing casters (2) two with locking brake	
	Multifunction cart with drawer	
	With holding frame for IV Fluids (3-4 hooks)	
	Dimension: (L x W x H) 850 - 870 x 500-520 x 1010 - 1030mm	
6	<b>VACUUM SUCTION, WALL REGULATOR WITH BOTTLE</b>	
	Suction Wall Bottle with holder	
	1600ml to 2000ml capacity plastic collection bottle	
	Autoclavable bottle	
	With secure lid that protects against overflow	
	With fitted tubing connector, at least 40cm in length	
	With bacterial filter at least five (5) pieces per set	
	With regulator securely fitted to NKTI existing central vacuum connectors	
	Regulator Range: 0 to 760 mm Hg Easy to manipulate knob and lever control	
	On/off selector use without changing original setting	
	With secure holder to be attached on the wall that will fit to existing wall suction lines	

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_