



**National Kidney and Transplant Institute**  
**Special Bids and Awards Committee**  
East Avenue, Quezon City 1100  
981-0300 / 981-0400 local 1157  
<http://www.nkti.gov.ph>

### **SUPPLEMENTAL BID BULLETIN NO. 21-114-3**

#### **Supply, Delivery, Testing and Commissioning of One (1) complete unit brand new Ultrasonic Surgical Aspirator System**

#### **Bid Reference IB No. 21-114**

This Supplemental Bid Bulletin No. 21-114-3 is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

<b>Particulars</b>	<b>Changes</b>
<b>Bid Data Sheet</b>	<b>Revised Bid Data Sheet</b>
<b>Item No. 4</b> Certificate for at least three (3) years warranty on parts, unit and service and preventive maintenance or manufacturer specifications	<b>Revised Item No. 4</b> Certificate for at least <b>five (5)</b> years warranty on parts, unit and service and preventive maintenance or manufacturer specifications
<b>Item No. 10</b> In case of frequent breakdown of the equipment of at least three times in a month within the first year of warranty period of three (3) years, we shall immediately provide a service unit of the same specification, model and brand within one (1) day from notification and new replacement unit shall be provided within thirty (30) calendar days. Otherwise, penalty shall be applied (actual income opportunity loss shall be charge to the supplier)	<b>Revised Item No. 10</b> In case of frequent breakdown of the equipment of at least three times in a month within the first year of warranty period of <b>five (5)</b> years, we shall immediately provide a service unit of the same specification, model and brand within one (1) day from notification and new replacement unit shall be provided within thirty (30) calendar days. Otherwise, penalty shall be applied ( <b>Penalty will be based on the number of cases done in a month</b> )
<b>Item No. 11</b> List of <b>COMPLETED</b> similar brand and similar functionality local installations by the winning bidder in at least two (2) hospitals or health or research facilities excluding NkTI with contact person and phone numbers	<b>Revised Item No. 11</b> List of <b>COMPLETED</b> similar brand and similar functionality local installations by the winning bidder in at least <b>one (1)</b> hospital or health or research facilities excluding NkTI with contact person and phone numbers
<b>Item No. 12</b> Certificate of Satisfactory Equipment Performance within the last three (3) years from at least two (2) local or domestic clients issued and signed by the Authorized Officer of the Operating Room Department or Section or the Head of Agency except NkTI with letterhead and contact number	<b>Revised Item No. 12</b> Certificate of <b>Satisfactory</b> Equipment Performance within the last <b>five (5) years</b> from at least two (2) local or domestic clients issued and signed by the Authorized Officer of the Operating Room Department or Section or the Head of Agency except NkTI with letterhead and contact number

<p><b>Item No. 13</b></p> <p>For current and past NKTl suppliers, the winning bidder shall submit notarized sworn statement that their average performance evaluation for the last five (5) years is at least Very Satisfactory and there is no negative feedback or issues related to the installation, and testing and commissioning, as well as post acceptance services including the consummated warranty, post warranty preventive and corrective maintenance of the supplied radiologic equipment</p>	<p><b><i>Revised Item No. 13</i></b></p> <p>For current and past NKTl suppliers, the winning bidder shall submit notarized sworn statement that their average performance evaluation for the last five (5) years is at least Very Satisfactory and there is no negative feedback or issues related to the installation, and testing and commissioning, as well as post acceptance services including the consummated warranty, post warranty preventive and corrective maintenance of the supplied <b>Operating Room</b> equipment</p>
<p>Technical Specifications</p>	<p><i>Please see attached Revised Technical Specifications</i></p>

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 6<sup>th</sup> day of September 2021 in Quezon City.

**VIOLETA M. VALDERRAMA, MD**  
**SBAC Chairman**



**National Kidney and Transplant Institute**  
**Special Bids and Awards Committee**  
East Avenue, Quezon City 1100  
981-0300 / 981-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## ***Section VII.***

# ***Revised Technical Specifications***

### **Supply, Delivery, Testing and Commissioning of One (1) complete unit brand new Ultrasonic Surgical Aspirator System**

***Instruction:***

*Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.*

<b>Item</b>	<b>Specification</b>	<b>Statement of Compliance</b>
<b>1</b>	<b>Console / Control Unit</b>	
	With the following capability via button controls;	
	a) Ultrasonic Aspirator	
	- Frequency Range: 23kHz to 36 kHz	
	- increment and decrement by at least 1%	
	- maximum adjustment of at least 100%	
	b) Irrigation	
	- with flow rate between 1 - 10 cc per minute	
	c) Suction	
	- with built in suction pump of at least 500 mmHg	
	- increment and decrement by at least, 2% via push button	

	d) Footswitch with ultrasonics/vibration function and irrigation or flushing function	
	e) OEM cart with container for suction bottles, with four (4) caster wheels with brakes	
2	<b>ACCESSORIES</b>	
	2.1 Handpiece/s a. Two (2) for open (either straight and/or angulated) b. Two (2) for laparoscopic	
	With one (1) sterilization case	
	2.2. Tip torquing set / Wrench set - one (1) set	
	a. Compatible with the respective offered frequencies	
	2.3 Tubings a. Ten (10) pieces per handpiece per frequency	
	2.4 Tip a. Ten (10) pieces tips per hand piece	
	2.5 Built-in monopolar capability a. At least ten (10) each b. Compatible with the respective offered frequencies	
	2.6 Contamination Guard / Contamination Filter / Filter a. At least ten (10) b. Compatible with the respective offered brand	
3	Power Supply: 220 +/- 10% VAC, Single Phase, 60Hz or auto-volt	
3	Appropriate ONLINE UPS	
4	Per manufacturer's specification with additional rated, branded UL-listed fused extension cord if equipment power cord is less than 6 meters	

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_