



National Kidney and Transplant Institute
Special Bids and Awards Committee
 East Avenue, Quezon City 1100
 981-0300 / 981-0400 local 1157
<http://www.nkti.gov.ph>

SUPPLEMENTAL BID BULLETIN NO. 21-117-1

Supply, Delivery, Testing and Commissioning of Four (4) units Automatic Lateral Tilting Bed

Bid Reference IB No. 21-117

This Supplemental Bid Bulletin No. 21-117-1 is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

Particulars	Changes
Items on Bid Data Sheet	Revised Items on Bid Data Sheet
Item No. 10 Certificate of Satisfactory Equipment Performance within the last three (3) years from at least two (2) local or domestic clients issued and signed by the Authorized Officer of the Operating Room Department or the Head of Agency except NkTI with letterhead and contact number	Revised Item No. 10 Certificate of Satisfactory Equipment Performance within the last three (3) years from at least two (2) local or domestic clients issued and signed by the Authorized Officer or the Head of Agency except NkTI with letterhead and contact number
Item No. 12 Certificate of guarantee that the bidder shall conduct in-house applications training by a qualified trainer certified by the Principal or Manufacturer with the machine for Nursing and Medical Staff with issuance of applications training certification	Revised Item No. 12 Certificate of guarantee that the bidder shall conduct in-house applications training by a qualified trainer certified by the Principal or Manufacturer with the machine for Medical and Paramedical Staff with issuance of applications training certification
Item No. 13 Certificate of guarantee issued by the bidder that in case of breakdown and on-going repair, the bidder shall provide a service unit with a similar unit	Revised Item No. 13 DELETED
Item No. 15 Certificate of guarantee that the bidder shall provide demonstration of the same unit offered in the bid within a non-extendable five (5) calendar days upon receipt of notification	Revised Item No. 15 Certificate of guarantee that the bidder shall provide physical demonstration of the same unit offered in the bid within a non-extendable five (5) calendar days upon receipt of notification
Schedule of Requirements	Please see attached Revised Schedule of Requirements
Technical Specifications	Please see attached Revised Technical Specification

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 26TH day of August 2021 in Quezon City.

VIOLETA M. VALDERRAMA, MD
SBAC Chairman



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VII.

Revised Technical Specifications

Supply, Delivery, Testing and Commissioning of Four (4) units Automatic Lateral Tilting Bed

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

Item	Specification	Statement of Compliance
	AUTOMATIC LATERAL TILTING BED	
	1. Bed Electrical Mechanism: Bed height and tilt, lateral tilt, thigh rest, calf rest, bed extension,	
	2. Bed Construction: Column Construction	
	3. Automatic Lateral Tilting: <ul style="list-style-type: none">• up to 30 °	
	4. External dimensions: <ul style="list-style-type: none">• (Length x Width) 215-220 x 105-110 cm	
	5. Internal dimension (max mattress dim.): <ul style="list-style-type: none">• 205-210 X 85-88 cm	
	6. Bed height adjustment: <ul style="list-style-type: none">• lowest height: 40 - 45 cm• Highest height: 80 - 85 cm	
	7. Bed extension (Foot with Extender): <ul style="list-style-type: none">• up to 25 cm	
	8. Bed tilt adjustment (Trendelenburg/Reverse Trendelenburg): <ul style="list-style-type: none">• 13°-15° / 16°-18°	

	<p>9. Backrest adjustment:</p> <ul style="list-style-type: none"> • 70°-75° 													
	<p>10. Thigh rest adjustment:</p> <ul style="list-style-type: none"> • 35°-40° 													
	<p>11. Power Supply:</p> <ul style="list-style-type: none"> • 220 +/- 10% VAC, Single Phase, 60Hz or AUTOVOLT 													
	<p>12. Position Control Button:</p> <table border="1"> <thead> <tr> <th>Purpose/Function</th> <th>Button</th> </tr> </thead> <tbody> <tr> <td>Orthopneic Chair / Cardiac Position</td> <td>One (1) Button</td> </tr> <tr> <td>Auto-contour</td> <td>One (1) Button</td> </tr> <tr> <td>Cardio-Pulmonary Resuscitation</td> <td></td> </tr> <tr> <td> <ul style="list-style-type: none"> • Electric • Manual </td> <td>One (1) Button Lever Type</td> </tr> <tr> <td>Emergency Trendelenburg</td> <td>One (1) Button</td> </tr> </tbody> </table>	Purpose/Function	Button	Orthopneic Chair / Cardiac Position	One (1) Button	Auto-contour	One (1) Button	Cardio-Pulmonary Resuscitation		<ul style="list-style-type: none"> • Electric • Manual 	One (1) Button Lever Type	Emergency Trendelenburg	One (1) Button	
Purpose/Function	Button													
Orthopneic Chair / Cardiac Position	One (1) Button													
Auto-contour	One (1) Button													
Cardio-Pulmonary Resuscitation														
<ul style="list-style-type: none"> • Electric • Manual 	One (1) Button Lever Type													
Emergency Trendelenburg	One (1) Button													
	<p>13. Battery Back-up:</p> <ul style="list-style-type: none"> • Minimum of 25 cycles 													
	<p>14. Safe working load (SWL):</p> <ul style="list-style-type: none"> • 250-300 kgs 													
	<p>15. Maximum patient weight:</p> <ul style="list-style-type: none"> • 185-200 kgs 													
	<p>16. Weight (basic equipment):</p> <ul style="list-style-type: none"> • 120-150kgs / 260-270 lbs. 													
	<p>17. Other Features</p>													
	<p>a. Mattress platform: Four (4) Sections</p>													
	<p>b. C-Arm and X-Ray compatible</p>													
	<p>c. Equipped with patient restraints holders</p>													
	<p>d. Powder Coated Metal Sheets</p>													
	<p>e. With removable mattress cover</p>													
	<p>f. Casters:</p> <ul style="list-style-type: none"> • Four (4) Antistatic Castors • Material: Polyurethane • Diameter: at least Five (5) inches 													

	<p>g. Brake System:</p> <ul style="list-style-type: none"> Centrally Operated Castors 	
	<p>h. Retractable 5th Castor wheel</p>	
	<p>i. Siderails:</p> <ul style="list-style-type: none"> Split Polymer Plastic Siderails Integrated Control LCD/LED (manufacturer's standard size) Multi-board in head siderails 	
	<p>j. Controls:</p> <ul style="list-style-type: none"> Integrated on the siderails Foot control for lateral tilt 	
	<p>k. Wall Bumpers:</p> <ul style="list-style-type: none"> Four (4) Horizontal corner bumpers 	
	<p>l. Weighing Scale:</p> <ul style="list-style-type: none"> Built-in Weighing Scale, in kilograms and in pounds 	
	<p>m. Angle Indicator:</p> <ul style="list-style-type: none"> Manufacturer Standards 	
	<p>n. Accessory Rail:</p> <ul style="list-style-type: none"> Available laterally 	
	<p>o. Night Light:</p> <ul style="list-style-type: none"> Manufacturer Standards 	
	<p>p. Alarms:</p> <ul style="list-style-type: none"> Bed Exit Alarm 	
	<p>q. Mattress:</p> <ul style="list-style-type: none"> Polyurethane foam Memory foam Anti-decubitus mattress with monolithic preventive effect against pressure sores Water resistant mattress cover Technology Bond Covers: Welding and Stapling (seamless) Max capacity: 230 kg-250kg Dimensions: <ul style="list-style-type: none"> Length: 205cm to 210cm Width: 85cm to 100cm Height or Thickness: 14cm to 18cm The mattress should fit the bed frame 	
	<p>18. Accessories (per bed)</p>	
	<p>a. Stabilizing Pads for lateral tilt (five (5) pieces/set/bed)</p>	
	<p>b. IV Pole:</p> <ul style="list-style-type: none"> At least two (2) pieces per bed Telescopic 	

	<ul style="list-style-type: none"> • With adjustable locking mechanism with four (4) chrome metal hooks • Length: Manufacturer Standard 	
	<p>c. Lifting Pole Assembly and hand Grip:</p> <ul style="list-style-type: none"> • Fixed at head-end, max load: 75kg- 80kg 	
	<p>d. Oxygen Bottle Holder:</p> <ul style="list-style-type: none"> • 304 Stainless Steel • Vertically 	
	<p>e. Bag Holder:</p> <ul style="list-style-type: none"> • Located laterally 	
	<p>f. Overload Protector</p>	
	<p>g. Bedside cabinet:</p> <ul style="list-style-type: none"> • same brand as procured bed) • as per manufacturer standard • ISO Standard or its equivalent 	
	<p>h. Overbed table:</p> <ul style="list-style-type: none"> • same brand as procured bed) • as per manufacturer standard • ISO Standard or its equivalent 	
	<p>i. Footstool:</p> <ul style="list-style-type: none"> • Single step • Made of 304 stainless steel • Non-slip rubber matting on top • Tubular legs with rubber tip on each leg 	
	<p>j. IV Stand:</p> <ul style="list-style-type: none"> • 304 Stainless Steel • Four (4) pivot hooks • Adjustable height knob type (standard height) • Caster: Five (5) wheel 	

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____