



**National Kidney and Transplant Institute**  
**Special Bids and Awards Committee**  
 East Avenue, Quezon City 1100  
 981-0300 / 981-0400 local 1157  
<http://www.nkti.gov.ph/>

**SUPPLEMENTAL BID BULLETIN NO. 21-118-1**

***Supply, Delivery, Installation and Testing of Two (2) units Bio Safety Cabinet, Class II***

**Bid Reference IB No. 21-118**

This Supplemental Bid Bulletin **No. 21-118-1** is being issued to clarify, modify and amend items/specifications in the Bidding Documents for the aforementioned project, to wit:

| Particulars   | Changes   |
|---|---|
| <b>Bid Data Sheet</b>   | <b>Revised Bid Data Sheet</b>   |
| <b>Item No. 12</b><br>Certificate of guarantee that the bidder shall conduct in-house/virtual applications training by a qualified trainer certified by the Principal or Manufacturer with the machine for Medical Oncology Staff with issuance of applications training certification  | <b>Revised Item No. 12</b><br>Certificate of guarantee that the bidder shall conduct in-house/virtual <b>orientation on the operation of the equipment for the end-user and basic troubleshooting for biomedical technician</b> by a qualified trainer certified by the Principal or Manufacturer with the machine with issuance of applications <b>orientation</b> /training certification |
| <b>Item No. 15</b><br>Certificate of guarantee that the bidder shall provide, in the order of priority, one of the following within a non-extendable five (5) calendar days upon receipt of notification:<br>1. a demo unit (same unit/model offered in the bid) or<br>2. arrange a site visit within NCR or<br>3. live virtual demonstration of the unit offered under recording | <b>Revised Item No. 15</b><br><b>Certificate of guarantee that the bidder shall provide physical demonstration of the same unit offered in the bid within a non-extendable five (5) calendar days upon receipt of notification</b>  |
| <b>Technical Specifications</b>   | <b>Revised Technical Specification</b>  |
| Size: 4 feet  | <b>Dimension:</b>   |
| Internal<br>Width: at least 44"<br>Depth: at least 22"<br>Height: at least 22"  | 1.b Internal: <b>as per manufacturer specifications</b>   |
|   | <b>Please see attached Revised Technical Specification</b>  |

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

**SUPPLEMENTAL BID BULLETIN NO. 21-118-1**

***Supply, Delivery, Installation and Testing of Two (2) units Bio Safety Cabinet,  
Class II***

For guidance and information of all concerned.

Issued this 3<sup>rd</sup> day of September 2021 in Quezon City.

**(SGD.) VIOLETA M. VALDERRAMA**  
**SBAC Chairman**



**National Kidney and Transplant Institute**  
**Special Bids and Awards Committee**  
 East Avenue, Quezon City 1100  
 981-0300 / 981-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## Section VII

### Revised Technical Specifications

**Instruction:**

*Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.*

| Item | Specification  | Statement of Compliance |
|------|--|-------------------------|
|      | <b>Supply, Delivery, Installation and Testing of Two (2) units Bio Safety Cabinet, Class II</b>          |                         |
| 1    | <b>Dimension:</b>  |                         |
|      | 1.a External<br>Width: at least 48” but not more than 60”<br>Depth: at least 30”<br>Height: at least 79” |                         |
|      | 1.b Internal: <b>as per manufacturer specifications</b>  |                         |
| 2    | <b>Electrical Requirement:</b><br>220 +/-10 tolerance or Auto-volt, 60Hz                                 |                         |
| 3    | <b>Other Specifications</b>  |                         |
|      | 3.a Microprocessor Controlled  |                         |
|      | 3.b With Ultra Low Penetration Air (ULPA) or High Efficiency Particulate Air (HEPA Filtration System)    |                         |
|      | 3.c With factory installed UV Light  |                         |
|      | 3.d With audible and visual alarms/monitoring  |                         |

|  |   |  |
|--|---|--|
|  | 3.e With white fluorescent/LED light  |  |
|  | 3.f With Safety lock-out protection/designated safety level height            |  |
|  | 3.g Single blower fan system  |  |
|  | 3.h Stainless steel chamber and work tray                                     |  |
|  | 3.i Ductless Cabinet  |  |
|  | 3.j With Support stand  |  |
|  | 3.k Fixtures-gas and vacuum port and internal power outlet                    |  |
|  | At least two (2) screw type stainless drain Plugs on both ends of the carrier |  |

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_