



**SUPPLEMENTAL BID BULLETIN NO. 21-124-1**

*Supply, Delivery, Testing and Commissioning of Ten (10) units Warmer Machine*

**Bid Reference IB No. 21-124**

This Supplemental Bid Bulletin No. 21-124-1 is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

Particulars	Changes
<p><b>Schedule of Requirement</b></p> <p>Within Sixty (60) calendar days</p>	<p><b>Revised Schedule of Requirement</b></p> <p>Within <b>Ninety (90)</b> calendar days</p>
<p><b>Technical Specifications</b></p> <p>Item No. 1 General Description</p> <p>a. Portable, with handle on machine, detachable from carrier/cart/holder</p> <p>b. Mobile, four (4) pivoting wheels</p> <p>c. With integrated clamp on machine for attachment to intravenous poles post</p> <p>d. With compartment on detachable carrier/cart/holder</p> <p>e. With readily available supply of interoperable existing blankets that can be used for pediatric and adult patients, with various configurations and sizes</p>	<p><b>Revised Technical Specifications</b></p> <p>Revised Item No. 1 General Description</p> <p>a. Portable, with handle on machine, detachable from carrier/cart</p> <p>b. Mobile, four (4) pivoting wheels</p> <p>c. With integrated clamp on machine for attachment to intravenous poles post</p> <p>d. With compartment on detachable carrier/cart</p> <p>e. With readily available supply of interoperable existing blankets that can be used for pediatric and adult patients, with various configurations and sizes</p>
<p>Item No. 4 Temperature</p> <p>With four (4) temperature settings or more</p> <p>a. Ambient</p> <p>b. Low</p> <p>c. Medium</p> <p>d. High</p>	<p>Revised Item No. 4 Temperature <b>(as per manufacturer's specifications)</b></p> <p>With four (4) temperature settings or more</p> <p>a. <b>Ambient/Room Air Temperature</b></p> <p>b. Low</p> <p>c. Medium</p> <p>d. High</p>
<p>Item No. 6 Heat Product and Transfer</p> <p>a. High Flow</p> <p>b. Convection type heat transfer</p> <p>c. Heater wattage +950 watts or its equivalent</p> <p>d. Hose: 1.5 meters length</p> <p style="padding-left: 20px;">d.1 Hose: Hose end temperature sensor or control or hose only</p> <p style="padding-left: 20px;">d.2 Corrugated type</p> <p style="padding-left: 20px;">d.3 Has mode of attachment to bed or stand</p> <p style="padding-left: 20px;">d.4 Readily available parts if damaged</p>	<p>Revised Item No. 6 Heat Product and Transfer</p> <p>a. High Flow</p> <p>b. Convection type heat transfer</p> <p>c. Heater wattage +950 watts or its equivalent</p> <p>d. Hose: <b>At least</b> 1.5 meters length</p> <p style="padding-left: 20px;">d.1 Hose: Hose end temperature sensor or control or hose only</p> <p style="padding-left: 20px;">d.2 Corrugated type</p> <p style="padding-left: 20px;">d.3 Has mode of attachment to bed or stand</p> <p style="padding-left: 20px;">d.4 Readily available parts if damaged</p>
<p>Item No. 7 Accessories</p>	<p>Revised Item No. 7 Accessories <b>per unit</b></p>

<b>Schedule of Requirements</b>	<b>Please see attached Revised Schedule of Requirements</b>
<b>Technical Specifications</b>	<b>Please see attached Revised Technical Specification</b>

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 17<sup>th</sup> day of September 2021 in Quezon City.

**VIOLETA M. VALDERRAMA, MD**  
**SBAC Chairman**



Republic of the Philippines  
National Kidney and Transplant Institute  
Bids and Awards Committee  
East Avenue, Quezon City 1100  
8981-0300 / 9881-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## Section VI.

### *Revised Schedule of Requirements*

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Qty./ UOM	Total	Delivered, Weeks/Months	Place of Delivery
1	Supply, Delivery, Testing and Commissioning of Ten (10) units Warmer Machine	1 lot		Within <b>Ninety (90)</b> calendar days	NKTI SMD

**Standard Payment Terms:** Payments shall be made promptly by NKTI, within sixty (60) calendar days from submission of complete documents *i.e.* invoice or claim by the Supplier.

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_



Republic of the Philippines  
National Kidney and Transplant Institute  
Bids and Awards Committee  
East Avenue, Quezon City 1100  
8981-0300 / 9881-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## ***Section VII.*** ***Technical Specifications***

### ***Supply, Delivery, Testing and Commissioning of Ten (10) units Warmer Machine***

***Instruction:***

*Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.*

<b>Item</b>	<b>Specification</b>	<b>Statement of Compliance</b>
	<b>PATIENT WARMER MACHINE (10 UNITS)</b>	
<b>1</b>	<b>General Description</b>	
	a. Portable, with handle on machine, detachable from carrier/cart	
	b. Mobile, four (4) pivoting wheels	
	c. With integrated clamp on machine for attachment to intravenous poles post	
	d. With compartment on detachable carrier/cart	
	e. With readily available supply of interoperable existing blankets that can be used for pediatric and adult patients, with various configurations and sizes	
<b>2</b>	<b>Power Supply: 220VAC +/-10%, Single Phase, 60Hz</b>	
<b>3</b>	<b>Filter</b>	
	a. With high efficiency filter of at least 0.3 µm filter	

4	<b>Temperature (as per manufacturer's specifications)</b>	
	With four (4) temperature settings or more	
	a. Ambient/Room Air Temperature	
	b. Low	
	c. Medium	
	d. High	
5	<b>Safety and Alarms</b>	
	a. Must have audible and visual alarms	
	b. With independent thermostat control for shut-off of heater when it exceeds at least 53°C	
	b.1 (+) 1°C above the selected temperature and/or above 45°C	
	b.2 (-) 1°C under the selected temperature and/or below	
6	<b>Heat Product and Transfer</b>	
	High Flow	
	Convection type heat transfer	
	Heater wattage +950 Watts or its equivalent	
	Hose: At least 1.5 meters length	
	d.1 Hose end temperature sensor or control or hose only	
	d.2 Corrugated type	
	d.3 Has mode of attachment to bed or stand	
	d.4 Readily available parts if damaged	
7	<b>Accessories per unit</b>	
	a. Ten (10) pieces full body warming blanket, adult	
	b. Ten (10) pieces full body warming blanket, pedia	

	c. Ten (10) pieces under body blanket, adult	
	d. Ten (10) pieces under body blanket, pedia	
	e. Ten (10) multi-position upper body blanket	
	<b>NOTE: With laser etching/engraving or similar permanent labeling method on handle or body of machine (NKT1 OR-RR 2021 and NUMBER OF MACHINE ex. No. 1, No. 2 and so on...)</b>	

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_