



National Kidney and Transplant Institute
Special Bids and Awards Committee
East Avenue, Quezon City 1100
981-0300 / 981-0400 local 1157
<http://www.nkti.gov.ph>

SUPPLEMENTAL BID BULLETIN NO. 21-126-1

Supply, Delivery Testing and Commissioning of Two (2) units Complete Combined/Dual Pneumatic/Shockwave and Ultrasonic Lithotripsy System

Bid Reference IB No. 21-126

This Supplemental Bid Bulletin No. 21-126-1 is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

| Particulars | Changes |
|---|---|
| Project Title Supply, Delivery Testing and Commissioning of Two (2) units Complete Combined Pneumatic and Ultrasonic Lithotripsy System | Revised Project Title Supply, Delivery Testing and Commissioning of Two (2) units Complete Combined/Dual Pneumatic/Shockwave and Ultrasonic Lithotripsy System |
| Special Conditions of the Contract | Revised Special Conditions of the Contract Additional Clause Clause 6 Liquidated Damages under Section 68 of Revised IRR of RA 9184 All contracts executed in accordance with the Act and this IRR shall contain a provision on liquidated damages which shall be payable by the contractor in case of breach thereof. For the procurement of Goods, Infrastructure Projects and Consulting Services, the amount of the liquidated damages shall be at least equal to one-tenth of one percent (0.001) of the cost of the unperformed portion for every day of delay. Once the cumulative amount of liquidated damages reaches ten percent (10%) of the amount of the contract, the Procuring Entity may rescind or terminate the contract, without prejudice to other courses of action and remedies available under the circumstances. |
| Schedule of Requirement Within Sixty (60) calendar days | Revised Schedule of Requirement Within Ninety (90) calendar days |
| Technical Specifications | Please see attached Revised Technical Specification |
| FINANCIAL COMPONENT ENVELOPE <input type="checkbox"/> (a) Original of duly signed and accomplished Financial Bid Form; <input type="checkbox"/> (b) Original of duly signed and accomplished Price Schedule(s); <input type="checkbox"/> (c) Original of duly signed and accomplished Price Schedule(s) | REVISED FINANCIAL COMPONENT ENVELOPE <input type="checkbox"/> (a) Original of duly signed and accomplished Financial Bid Form; <input type="checkbox"/> (b) Original of duly signed and accomplished Price Schedule(s); b.1 Price Schedule for Goods Offered from Abroad or |

| | |
|---|--|
| b.1 Price Schedule for Goods Offered from Abroad or b.2 Price Schedule for Goods from within the Philippines, and b.3 Annex "A" Bid Proposal Form including cost breakdown for accessories | b.2 Price Schedule for Goods from within the Philippines, and b.3 Annex "A" Bid Proposal Form including cost breakdown for accessories |
| Bid Proposal Form | Please see attached Revised Bid Proposal Form |

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 27th day of September 2021 in Quezon City.

VIOLETA M. VALDERRAMA, MD
SBAC Chairman

REVISED BID PROPOSAL FORM

**IB No. 21-126: Supply, Delivery Testing and Commissioning of Two (2) units Complete Combined/Dual
Pneumatic/Shockwave and Ultrasonic Lithotripsy System**

Opening of Bids: _____
Claver Ramos Auditorium, G/F NKTl Main Building

| Item No. | Particulars | QTY/UOM | Approved Budget Cost | | Supplier | BID PROPOSAL | | Remarks |
|--------------------|--|---------|----------------------|---------------------|----------|--------------|--------------|---------|
| | | | Unit Cost | Total Amount | | Unit Cost | Total Amount | |
| 1 | Supply, Delivery Testing and Commissioning of Complete Combined/Dual Pneumatic/Shockwave and Ultrasonic Lithotripsy System | 2 units | 4,400,000.00 | 8,800,000.00 | | | | |
| GRAND TOTAL | | | | 8,800,000.00 | | | | |

This Bid Proposal shall be accompanied by cost breakdown of accessories

Printed Name of Company

Date

Signature

Address

Telephone Number

Printed Name and Designation



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VI.

Revised Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

| Item Number | Description | Qty./ UOM | Total | Delivered, Weeks/Months | Place of Delivery |
|-------------|---|-----------|-------|----------------------------------|-------------------|
| 1 | Supply, Delivery Testing and Commissioning of Two (2) units Combined/dual Pneumatic/shockwave and Ultrasonic Lithotripsy System | 1 lot | | Within Ninety (90) calendar days | NKTI SMD |

Standard Payment Terms: Payments shall be made promptly by NKTI, within sixty (60) days from submission of complete documents *i.e.* invoice or claim by the Supplier.

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VII.

Revised Technical Specifications

Supply, Delivery Testing and Commissioning of Two (2) units Complete Combined/Dual Pneumatic/Shockwave and Ultrasonic Lithotripsy System

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

| Item | Specification | Statement of Compliance |
|-------------|--|--------------------------------|
| 1 | Lithotripsy Unit | |
| | a. With pneumatic and ultrasonic capability | |
| | b. Capable of individual (pneumatic only or ultrasonic only) or dual function | |
| | c. With suction function | |
| 2 | Handpiece | |
| | a. With pneumatic and ultrasonic function may be single or separate per capability | |

| | | |
|----------|--|--|
| | b. With appropriate adapters | |
| | c. With nose piece seal when applicable | |
| | d. With individual case for handpiece and accessories | |
| 3 | Suction Probes | |
| | a. Suction probes should be available when applicable | |
| | b. Suction tubing at least 35 cm length | |
| | c. With individual case for handpiece and accessories | |
| 4 | Suction Pump | |
| | a. Adjustable pressure, built-in or standalone within machine parameters | |
| 5 | Cart | |
| | a. Medical grade equipment cart for unit | |
| | b. OEM | |
| | c. With brakes on at least two casters | |
| 6 | Consumables | |
| | 6.1 Probes (two (2) pieces for each type, a total of four (4) pieces probes) | |
| | a. Sterilizable | |
| | b. Reusable probes with pneumatic function with the following dimensions: | |
| | b.1 Pneumatic Probe (two (2) pieces per unit) | |
| | b.1.1 Diameter: 1mm to 3mm | |
| | b.1.2 Length: more than 30cm | |
| | c. Reusable probes with ultrasonic function good for at least five (5) uses with the following dimensions: | |
| | c.1 Ultrasonic Probe (two (2) pieces per unit) | |
| | c.1.1 Diameter: 1.3mm to 4mm | |
| | c.1.2 Length: more than 30cm | |
| | c.1.3 Able to handle at least 20,000 Hz | |

| | | |
|---|--|--|
| 7 | POWER SUPPLY | |
| | d. With at least one (1) online UPS per unit with bracket or mount to equipment for UPS | |
| | e. As per Manufacturer's specification with additional rated, branded UL-listed fused extension cord if equipment power cord is less than six (6) meters | |
| | | |

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____