



SUPPLEMENTAL BID BULLETIN NO. 21-131-1

One (1) Lot Supply, Delivery, Testing and Commissioning of One (1) unit Cordless Battery Powered Surgical Oscillating Saw with Accessories and One (1) unit Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories

Bid Reference IB No. 21-131

This Supplemental Bid Bulletin No. 21-131-1 is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

Particulars	Changes
Schedule of Requirements	Revised Schedule of Requirements
Within sixty (60) calendar days	Within Ninety (90) calendar days <i>Please see attached Revised Schedule of Requirements</i>
Technical Specifications	Revised Technical Specifications
Cordless Battery Powered Surgical Oscillating Saw with Accessories	Cordless Battery Powered Surgical Oscillating Saw with Accessories
1.3 Accessories:	1.3 Accessories:
1.3.1 Sterilization Case – One (1) piece	1.3.1 Sterilization Case, with integrated permanent germ barrier system made of PTFE reusable filter for at least 5,000 sterilization cycles – One (1) piece with two (2) reusable filter
1.3.6 Charging unit for batteries when applicable one (1) piece	1.3.6 Charging unit - one (1) piece with four (4) charging day/dock
Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories	Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories
2.4 Accessories	2.4 Accessories
2.4.1 Sterilization Case – One (1) piece	2.4.1 Sterilization Case, with integrated permanent germ barrier system made of PTFE reusable filter for at least 5,000 sterilization cycles – One (1) piece with two (2) pieces reusable filter
2.4.7 Charging unit for batteries when applicable one (1) piece	2.4.7 Charging unit - one (1) piece with four (4) charging day/dock
	Please see attached Revised Technical Specifications

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 7th day of October 2021 in Quezon City.

VIOLETA M. VALDERRAMA, MD
SBAC Chairman



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VI.

Revised Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Qty./ UOM	Total	Delivered, Weeks/Months	Place of Delivery
1	Supply, Delivery, Testing and Commissioning of One (1) unit Cordless Battery Powered Surgical Oscillating Saw with Accessories and One (1) unit Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories	One (1) lot		Within Ninety (90) calendar days	NKTI SMD

Standard Payment Terms: Payments shall be made promptly by NKTI, within sixty (60) calendar days from submission of complete documents *i.e.* invoice or claim by the Supplier.

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VII.

Revised Technical Specifications

One (1) Lot Supply, Delivery, Testing and Commissioning of One (1) unit Cordless Battery Powered Surgical Oscillating Saw with Accessories and One (1) unit Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against EACH of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

Item	Specification	Statement of Compliance
1	Cordless Battery Powered Surgical Oscillating Saw with Accessories	
	1.1 Processing:	
	1.1.1 Washable - except for internal components	
	1.1.2 Autoclavable - all parts except for batteries/internal components	
	1.2 Features:	
	1.2.1 Oscillating type	
	1.2.2 Reusable	
	1.2.3 Handheld, Cordless	
	1.2.4 Battery powered - with indicator/notification for energy/power level (at least 2 -3 hours operating time)	

	1.3 Accessories:	
	1.3.1 Sterilization Case, with integrated permanent germ barrier system made of PTFE reusable filter for at least 5,000 sterilization cycles – One (1) piece with two (2) pieces reusable filter	
	1.3.2 At least ten (10) pieces blades - adult	
	1.3.3 At least five (5) pieces blades - pediatric	
	1.3.4 At least two (2) batteries	
	1.3.5 Assembling tools when applicable	
	1.3.6 Charging unit - one (1) piece with four (4) charging day/dock	
2	Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories	
	2.1 Processing:	
	2.1.1 Washable - except for internal components	
	2.1.2 Autoclavable - all parts except for batteries/internal components	
	2.2 Features:	
	2.2.1 Reciprocating type	
	2.2.2 Reusable	
	2.2.3 Handheld, Cordless	
	2.2.4 Battery powered - with indicator/notification for energy/power level (at least 2 -3 hours operating time)	
	2.3 Protection/Safety:	
	2.3.1 With adult and pediatric sternum protector one (1) each	
	2.4 Accessories	
	2.4.1 Sterilization Case, with integrated permanent germ barrier system made of PTFE reusable filter for at least 5,000 sterilization cycles – One (1) piece with two (2) pieces reusable filter	
	2.4.2 At least ten (10) pieces sternum blades - adult	
	2.4.3 At least five (5) pieces sternum blades - pediatric	

	2.4.4 At least two (2) sternum guard	
	2.4.5 At least two (2) batteries	
	2.4.6 Assembling tools when applicable	
	2.4.7 Charging unit - one (1) piece with four (4) charging day/dock	

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____