



**National Kidney and Transplant Institute**  
**Special Bids and Awards Committee**  
 East Avenue, Quezon City 1100  
 981-0300 / 981-0400 local 1157  
<http://www.nkti.gov.ph>

**SUPPLEMENTAL BID BULLETIN NO. 21-131-3**

*One (1) Lot Supply, Delivery, Testing and Commissioning of One (1) unit Cordless Battery Powered Surgical Oscillating Saw with Accessories and One (1) unit Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories*

**Bid Reference IB No. 21-131**

This Supplemental Bid Bulletin **No. 21-131-3** is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

Particulars	Changes
<b>PROJECT TITLE</b>	<b>REVISED PROJECT TITLE</b>
One (1) Lot Supply, Delivery, Testing and Commissioning of One (1) unit Cordless Battery Powered Surgical Oscillating Saw with Accessories and One (1) unit Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories	Supply, Delivery, Testing and Commissioning of One (1) unit Cordless Battery Powered Surgical Oscillating Saw with Accessories and One (1) unit Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories (Line Item)
<b>APPROVED BUDGET FOR THE CONTRACT</b>	<b>REVISED APPROVED BUDGET FOR THE CONTRACT</b>
Four Million Pesos only (Php 4,000,000.00)	Three Million Seven Hundred Thousand Pesos (Php 3,700,000.00)
<b>BID DATA SHEET</b>	<b>BID DATA SHEET</b>
Clause 19.3 All items to be grouped together to form one complete Lot that will be awarded to one bidder to form one complete contract	Revised Clause 19.3 One Project having several items, which shall be awarded as separate contracts per item
<b>TECHNICAL SPECIFICATIONS</b>	<b>REVISED TECHNICAL SPECIFICATIONS</b>
Cordless Battery Powered Surgical Oscillating Saw with Accessories	Cordless Battery Powered Surgical Oscillating Saw with Accessories
1.3 Accessories:	1.3 Brand New Accessories:
1.3.1 Sterilization Case, with integrated permanent germ barrier system made of PTFE reusable filter for at least 5,000 sterilization cycles – One (1) piece with two (2) reusable filter	1.3.1 Sterilization Case – One (1) piece as per manufacturer’s specifications
1.3.6 Charging unit - one (1) piece with four (4) charging day/dock	1.3.6 Charging unit - one (1) piece as per manufacturer’s specifications
Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories	Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories
2.4 Accessories	2.4 Brand New Accessories
2.4.1 Sterilization Case, with integrated permanent germ barrier system made of PTFE reusable filter for at least 5,000 sterilization cycles – One (1) piece with two (2) pieces reusable filter	2.4.1 Sterilization Case – One (1) piece as per manufacturer’s specifications

2.4.7 Charging unit - one (1) piece with four (4) charging day/dock					2.4.7 Charging unit - one (1) piece as per manufacturer's specifications				
					<b>Please see attached Revised Technical Specifications</b>				
<b>BID PROPOSAL FORM</b>					<b>REVISED BID PROPOSAL FORM</b>				
Item No.	Particulars	QTY/ UOM	Approved Budget Cost		Item No.	Particulars	QTY/ UOM	Approved Budget Cost	
			Unit Cost	Total Amount				Unit Cost	Total Amount
1	Cordless Battery Powered Surgical Oscillating Saw with Accessories	1 unit	1,500,000.00	1,500,000.00	1	Cordless Battery Powered Surgical Oscillating Saw with Accessories	1 unit	1,850,000.00	1,850,000.00
2	Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories	1 unit	2,500,000.00	2,500,000.00	2	Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories	1 unit	1,850,000.00	1,850,000.00
<b>GRAND TOTAL</b>				<b>4,000,000.00</b>	<b>GRAND TOTAL</b>				<b>3,700,000.00</b>
					<b>Please see attached Revised Bid Proposal Form</b>				

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 20<sup>th</sup> day of October 2021 in Quezon City.

**VIOLETA M. VALDERRAMA, MD**  
**SBAC Chairman**

**REVISED BID PROPOSAL FORM**

**IB No. 21-131: Supply, Delivery, Testing and Commissioning of One (1) unit Cordless Battery Powered Surgical Oscillating Saw with Accessories and One (1) unit Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories (Line Item)**

Opening of Bids: \_\_\_\_\_

**BAC Conference Room, G/F NKT Main Building**

Item No.	Particulars	QTY/ UOM	Approved Budget Cost		Supplier	BID PROPOSAL		Remarks
			Unit Cost	Total Amount		Unit Cost	Total Amount	
1	Cordless Battery Powered Surgical Oscillating Saw with Accessories	1 unit	1,850,000.00	1,850,000.00				
2	Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories	1 unit	1,850,000.00	1,850,000.00				
<b>GRAND TOTAL</b>				<b>3,700,000.00</b>				

This Bid Proposal shall be accompanied by cost breakdown of accessories

\_\_\_\_\_  
Printed Name of Company

\_\_\_\_\_  
Date

\_\_\_\_\_  
Signature

\_\_\_\_\_  
Address

\_\_\_\_\_  
Telephone Number

\_\_\_\_\_  
Printed Name and Designation



Republic of the Philippines  
National Kidney and Transplant Institute  
Bids and Awards Committee  
East Avenue, Quezon City 1100  
8981-0300 / 9881-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## ***Section VII.***

### ***Second Revised Technical Specifications***

#### ***Supply, Delivery, Testing and Commissioning of One (1) unit Cordless Battery Powered Surgical Oscillating Saw with Accessories and One (1) unit Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories (Line Item)***

***Instruction:***

*Bidders must state in the column provided either “Comply” or “Not Comply” against EACH of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.*

<b>Item</b>	<b>Specification</b>	<b>Statement of Compliance</b>
1	<b>Cordless Battery Powered Surgical Oscillating Saw with Accessories</b>	
	<b>1.1 Processing:</b>	
	1.1.1 Washable - except for internal components	
	1.1.2 Autoclavable - all parts except for batteries/internal components	
	<b>1.2 Features:</b>	
	1.2.1 Oscillating type	
	1.2.2 Reusable	
	1.2.3 Handheld, Cordless	
	1.2.4 Battery powered - with indicator/notification for energy/power level (at least 2 -3 hours operating time)	

	<b>1.3 Brand New Accessories:</b>	
	1.3.1 Sterilization Case – One (1) piece as per manufacturer’s specifications	
	1.3.2 At least ten (10) pieces blades - adult	
	1.3.3 At least five (5) pieces blades - pediatric	
	1.3.4 At least two (2) batteries	
	1.3.5 Assembling tools when applicable	
	1.3.6 Charging unit - one (1) piece as per manufacturer’s specifications	
2	<b>Cordless Battery Powered Surgical Reciprocating Sternal Saw with Accessories</b>	
	<b>2.1 Processing:</b>	
	2.1.1 Washable - except for internal components	
	2.1.2 Autoclavable - all parts except for batteries/internal components	
	<b>2.2 Features:</b>	
	2.2.1 Reciprocating type	
	2.2.2 Reusable	
	2.2.3 Handheld, Cordless	
	2.2.4 Battery powered - with indicator/notification for energy/power level (at least 2 -3 hours operating time)	
	<b>2.3 Protection/Safety:</b>	
	2.3.1 With adult and pediatric sternum protector one (1) each	
	<b>2.4 Brand New Accessories</b>	
	2.4.1 Sterilization Case – One (1) piece as per manufacturer’s specifications	
	2.4.2 At least ten (10) pieces sternum blades - adult	
	2.4.3 At least five (5) pieces sternum blades - pediatric	
	2.4.4 At least two (2) sternum guard	
	2.4.5 At least two (2) batteries	

	2.4.6 Assembling tools when applicable	
--	--	--

	2.4.7 Charging unit - one (1) piece as per manufacturer's specifications	
--	--	--

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_