



National Kidney and Transplant Institute
Special Bids and Awards Committee
East Avenue, Quezon City 1100
981-0300 / 981-0400 local 1157
<http://www.nkti.gov.ph>

SUPPLEMENTAL BID BULLETIN NO. 21-133-1

Supply, Delivery, Testing and Commissioning of Two (2) units Brain Function Consciousness Monitor and Eight (8) units Enteral Feeding Pump (Line Bidding)

Bid Reference IB No. 21-133

This Supplemental Bid Bulletin No. 21-133-1 is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

Particulars	Changes
Bid Data Sheet	Revised Bid Data Sheet
Clause 21.2 (item 11) For current and past NKTI suppliers, the winning bidder shall submit notarized sworn statement that their average performance evaluation for the last five (5) years is at least Very Satisfactory and there is no negative feedback or issues related to the installation, and testing and commissioning, as well as post acceptance services including the consummated warranty, post warranty preventive and corrective maintenance of the supplied Medical equipment	Revised Clause 21.2 (item 11) Revised item No. 2: For current and past NKTI suppliers, the winning bidder shall submit notarized sworn statement that their average performance evaluation for the last five (5) years is at least Very Satisfactory and there is no negative feedback or issues related to the installation, and testing and commissioning, as well as post acceptance services including the consummated warranty, post warranty preventive and corrective maintenance of the supplied equipment
Schedule of Requirements	Revised Schedule of Requirements
Delivered, Weeks/Months: Within sixty (60) calendar days	Delivered, Weeks/Months: Within Ninety (90) calendar days <i>Please see attached Revised Schedule of Requirements</i>
Technical Specifications	Revised Technical Specifications

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 14th day of October 2021 in Quezon City.

VIOLETA M. VALDERRAMA, MD
SBAC Chairman



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VI.

Revised Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Qty./ UOM	Total	Delivered, Weeks/Months	Place of Delivery
1	Brain Function Consciousness Monitor	2 units		Within Ninety (90) calendar days	NKTI SMD
2	Enteral Feeding Pump	8 units		Within Ninety (90) calendar days	NKTI SMD

Standard Payment Terms: Payments shall be made promptly by NKTI, within sixty (60) calendar days from submission of complete documents *i.e.* invoice or claim by the Supplier.

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VII.

Revised Technical Specifications

Supply, Delivery, Testing and Commissioning of Two (2) units Brain Function Consciousness Monitor and Eight (8) units Enteral Feeding Pump (Line Bidding)

Instruction: <i>Bidders must state in the column provided either “Comply” or “Not Comply” against EACH of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.</i>		
Item	Specification	Statement of Compliance
1	BRAIN FUNCTION CONCIOUSNESS MONITOR (2 UNITS)	
	1.1 Can process EEG information and calculates a number between 0 and 100	
	1.2 The system can perform self-tests to ensure that the monitor and its components are functioning properly and that impedance levels of patient sensors are within acceptable limits	
	1.3 Composition:	
	1.3.1 Monitor, processor and interfaced cable	
	1.4 Has an IV pole clamp	
	1.5 Power Supply:	
	1.5.1 220VAC +/-10%, 60Hz, Single Phase or AUTOVOLT	

	1.6 Interface Cable with Processor:	
	1.6.1 Interoperable with existing brain function/consciousness monitor sensor	
	1.7 Data Gathering:	
	1.7.1 Can gather data over the last 15 to 30 seconds of EEG recording and updated every second	
	1.7.2 Capable to display graphical trend	
	1.8 Battery:	
	1.8.1 Rechargeable	
	1.9 Monitor:	
	1.9.1 Two (2) channel monitor	
	1.9.2 Touch Screen	
	1.9.2 At least One (1) USB port	
	1.10 Accessories	
	10.1 Sensors – Twenty-Five (25) pieces per unit	
2	ENTRAL FEEDING PUMP (8 UNITS)	
	2.1 Applicable for both continuous and intermittent feeding	
	2.2 With Flushing capacity	
	2.3 User-Friendly interface with adjustable setting and readable display:	
	2.3.1 Volume to be delivered in ml increments of 1ml	
	2.3.2 Time of feeding in hour increments of 1 hour	
	2.3.3 Feed/flush rate in ml/hour increments of 1ml	
	2.4 Accurately delivers volume/hour as set within 5% margin of error	
	2.5 Flush rate within 1ml to 600ml	
	2.6 Flow rate at least 1ml – 400 ml/hour (1 ml increment)	

	2.7 Has priming function (auto priming)	
	2.8 Accessories	
	2.8.1 Enteral Feeding bag - Fifty (50) pcs per unit	
	Note: Submit copy of brochure including complete technical data sheet	

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____