



**National Kidney and Transplant Institute**  
**Special Bids and Awards Committee**  
East Avenue, Quezon City 1100  
981-0300 / 981-0400 local 1157  
<http://www.nkti.gov.ph/>

**SUPPLEMENTAL BID BULLETIN NO. 21-136-1**

***Supply, Delivery, Testing and Commissioning of Three (3) units Vital Signs Monitor with Wireless Capability and Bar Code Reader***

**Bid Reference IB No. 21-136**

This Supplemental Bid Bulletin **No. 21-136-1** is being issued to clarify, modify and amend items/specifications in the Bidding Documents for the aforementioned project, to wit:

<b>Particulars</b>	<b>Changes</b>
<b>Technical Specifications</b>	<b>Revised Technical Specifications</b>
<b><i>Item letter (I)</i></b>  With Accessory Power Management - With external reserved battery to ensure performance even unplugged for a long time	<b><i>Revised Item letter (I)</i></b>  With external reserve battery (integrated or separate)  <b>Please see attached Revised Technical Specification</b>

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 18<sup>th</sup> day of October 2021 in Quezon City.

**VIOLETA M. VALDERRAMA**  
**SBAC Chairman**



**National Kidney and Transplant Institute**  
**Special Bids and Awards Committee**  
 East Avenue, Quezon City 1100  
 981-0300 / 981-0400 local 1156/1186  
<http://www.nkti.gov.ph/>

## Section VII

### Revised Technical Specifications

#### Supply, Delivery, Testing and Commissioning of Three (3) units Vital Signs Monitor with Wireless Capability and Bar Code Reader

**Instruction:**

*Bidders must state in the column provided either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.*

Item	Specification	Statement of Compliance
<b>1</b>	<b>VITAL SIGNS MONITOR WITH WIRELESS CAPABILITY AND BAR CODE READER</b>	
	Capable of Spot Check, Monitoring, and triaging purpose.	
	a. Parameters: Blood pressure, Heart/Pulse Rate, Respiratory Rate, Oxygen Saturation, and Temperature (Tympanic)	
	b. Includes wide range of BP cuff: Infant, pedia, small adult, adult and large adult patient (2 sets)	
	c. Includes reusable two (2) adult and One (1) pediatric Oxygen saturation reader.	
	d. Records VS and has review setting (individual patient or list).	
	e. Configurable alarm parameters.	
	f. with Barcode reader for patient identifier and to synchronize with the system.	
	g. EMR connectivity ready (at least HL7)	
	h. Wirelessly transmit vital signs from the bedside to the EMR.	

	i. Software for connection to EMR must be provided by the winning bidder.	
	j. Battery: Li-ion rechargeable battery, with battery gauge indicator.	
	k. With medical grade trolley powder coated with basket/compartiment which holds the accessories with five (5) caster wheels with at least two (2) wheels with brake.	
	l. <b>With external reserve battery (integrated or separate)</b>	

**Conforme:**

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_