



National Kidney and Transplant Institute
Special Bids and Awards Committee
East Avenue, Quezon City 1100
981-0300 / 981-0400 local 1157
<http://www.nkti.gov.ph>

SUPPLEMENTAL BID BULLETIN NO. 21-137-1

One (1) Lot Supply, Delivery, Testing and Commissioning of Two (2) units Bipolar Cautery Machine and Endoscope Tower, with Two (2) Bipolar Telescopes and Resectoscope sets (with Case and Cleaning Materials: Brush and Oil) and Various Electrodes

Bid Reference IB No. 21-137

This Supplemental Bid Bulletin **No. 21-137-1** is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

Particulars	Changes
Schedule of Requirements	Revised Schedule of Requirements
Delivered, Weeks/Months: Within Sixty (60) calendar days	Delivered, Weeks/Months: Within Ninety (90) calendar days <i>Please see attached Revised Schedule of Requirements</i>
Technical Specifications	Revised Technical Specifications

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 18th day of October 2021 in Quezon City.

VIOLETA M. VALDERRAMA, MD
SBAC Chairman



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VI.

Revised Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Qty./ UOM	Total	Delivered, Weeks/Months	Place of Delivery
1	Supply, Delivery, Testing and Commissioning of Two (2) units Bipolar Cautery Machine and Endoscope Tower, with Two (2) Bipolar Telescopes and Resectoscope sets (with Case and Cleaning Materials: Brush and Oil) and Various Electrodes	One (1) lot		Within Ninety (90) calendar days	NKTI SMD

Standard Payment Terms: Payments shall be made promptly by NKTI, within sixty (60) calendar days from submission of complete documents *i.e.* invoice or claim by the Supplier.

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VII.

Revised Technical Specifications

One (1) Lot Supply, Delivery, Testing and Commissioning of Two (2) units Bipolar Cautery Machine and Endoscope Tower, with Two (2) Bipolar Telescopes and Resectoscope sets (with Case and Cleaning Materials: Brush and Oil) and Various Electrodes

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against EACH of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

Item	Specification	Statement of Compliance
1	ENDOSCOPY TOWER (2) Units	
	1.1 Monitor	
	1.1.1 Medical Grade Monitor, panel size of at least 31 inches with at least 3840 x 2160 resolution	
	1.1.2 With image processing software or hardware (camera based)	
	1.1.3 Touchscreen or soft touch button interface	

	1.2 Lightsource	
	1.2.1 Xenon short-arc lamp or LED White Light	
	1.2.2 Adjustable brightness, white light available	
	1.2.3 Life Span: At least 10,000 hours	
	1.3 Lightcable	
	1.3.1 At least three (3) pieces per unit with at least three (3) meters length appropriate light guide cable with adapters	
	1.3.2 Processable for plasma sterilization	
	1.4 Camera Control Unit with light source	
	1.4.1 2D processing camera control unit with built-in or separate light source	
	1.4.2 With electronic zoom and optical digital observation	
	1.4.3 With LED light illumination system	
	1.4.4 Resolution: At least 3840 x 2160 resolution	
	1.5 Camera Head	
	1.5.1 3 CMOS quality or equivalent Camera head with pick up system, focal length of f=15.9 to 31.3mm, with electronic shutter and zoom functions	
	1.5.2 Autoclavable or sterilizable	
	1.5.3 With hard case and foam inserts or its equivalent	
	1.6 Endoscope/Telescope	
	1.6.1 12-degrees (4mm Outer Diameter) - one (1) piece per unit	
	1.6.2 Autoclavable	
	1.6.3 With hard case and foam inserts or its equivalent	
	1.7 Recording System	
	1.7.1 Full Medical Grade HD recorder, with still and video images recording - must have no degradation in image quality	
	1.7.2 Network capability thru Gigabit LAN port or USB port	
	1.8 Storage	

	1.8.1 At least two (2) pieces of 4TB external hard drive compatible with provided recording unit	
	1.9 Cart/Trolley	
	1.9.1 Medical Grade Heavy duty cart for holding all machine components in one place 1.9.2 With ports for power or per manufacturer's specification; additional rated, branded UL-listed fused extension cord with four (4) heavy duty casters with brakes	
	1.10 UPS	
	1.10.1 Online UPS Specific to units, one (1) per unit	
	1.11 Power Supply	
	1.11.1 voltage: 100-120V/220-240V or AUTOVOLT	
2	Resectoscope Set Bipolar (2 sets per unit)	
	2.1 Outer Sheath	
	2.1.1 Resectoscope Sheath, 26 Fr., oblique beak	
	2.2 Inner Sheath	
	2.2.1 Rotating inner sheath, with ceramic insulation, for use with Standard obturator, for 24/26 Fr. sheath	
	2.3 Working Element	
	2.3.1 Passive or Electromechanical, with working element	
	2.4 Bipolar/HF Cable	
	2.4.1 HF-cable, bipolar, 4m, compatible with generator and saline resectoscope	
	2.5 With hard case and foam inserts or its equivalent	
3	Endoscope/Telescope (2 per unit)	
	3.1 Telescope 12 degrees, enlarged view 4mm, autoclavable, fiber optic light transmission incorporated, color code: black	
	3.2 With hard case and foam inserts or its equivalent	

4	Electrosurgical Generator Set (1 per unit)	
	4.1 Electrosurgical Generator Set	
	4.1.1 Capable of Saline bipolar resection, loop resection, vaporization, and transurethral enucleation	
	4.1.2 With Monopolar/Bipolar Frequency Functions	
	4.1.3 Voltage: 100-120V/220-240V, AUTOVOLT	
	4.1.4 With monopolar sockets (2 sockets), 1 bipolar socket, 1 all universal grounding type socket	
	4.1.5 One (1) neutral electrode socket	
	4.1.6 With capability for open, lap, and endoscopic surgery procedures	
	4.1.7 With tissue-adaptive power output	
5	Bipolar Electrode Set (Compatible with Electrosurgical Set)	
	5.1 HF-resection electrode, loop, 24 Fr., 0.2 Fr. wire, medium, 12 degrees, sterile, single use, at least 60 pcs, for saline	
	5.2 HF-resection electrode, button, 24 Fr., 12 to 30 degrees, sterile, single use, at least 48pcs, for saline	
	5.3 HF-resection electrode, roller, 24-28 Fr., 12 to 30 degrees, sterile, single use, at least 12 pcs, for saline	
	5.4 HF-resection electrode, 45 degree needle, 24-28 Fr., 12 to 30 degrees, sterile, single use, at least 12pcs, for saline	
	5.5 HF-resection electrode, Angled Loop, 12 to 30 degrees, sterile, single use, at least 10pcs, for saline	
Conforme:		
Name: _____		
Legal Capacity: _____		
Signature: _____		
Duly authorized to sign the Bid for and behalf of: _____		