



National Kidney and Transplant Institute
Special Bids and Awards Committee
East Avenue, Quezon City 1100
981-0300 / 981-0400 local 1157
<http://www.nkti.gov.ph>

SUPPLEMENTAL BID BULLETIN NO. 21-138-1

***One (1) Lot Supply, Delivery, Testing and Commissioning of Various Endoscopes,
Interoperable with Existing Equipment***

Bid Reference IB No. 21-138

This Supplemental Bid Bulletin No. 21-138-1 is being issued to clarify, modify and amend items/specifications in the Bidding Documents and to address key issues raised during the Pre-Bid Conference for the aforementioned project, to wit:

Particulars	Changes
Schedule of Requirements	Revised Schedule of Requirements
Delivered, Weeks/Months: Within Sixty (60) calendar days	Delivered, Weeks/Months: Within Ninety (90) calendar days <i>Please see attached Revised Schedule of Requirements</i>
Technical Specifications	Revised Technical Specifications

This Supplemental Bid Bulletin including Annexes, if any, shall form part of the Bid Documents. Any provisions in the Bid Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 18th day of October 2021 in Quezon City.

VIOLETA M. VALDERRAMA, MD
SBAC Chairman



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VI.

Revised Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Qty./ UOM	Total	Delivered, Weeks/Months	Place of Delivery
1	Supply, Delivery, Testing and Commissioning of Various Endoscopes, Interoperable with Existing Equipment	One (1) lot		Within Ninety (90) calendar days	NKTI SMD

Standard Payment Terms: Payments shall be made promptly by NKTI, within sixty (60) calendar days from submission of complete documents *i.e.* invoice or claim by the Supplier.

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____



Republic of the Philippines
National Kidney and Transplant Institute
Bids and Awards Committee
East Avenue, Quezon City 1100
8981-0300 / 9881-0400 local 1156/1186
<http://www.nkti.gov.ph/>

Section VII.

Revised Technical Specifications

One (1) Lot Supply, Delivery, Testing and Commissioning of Various Endoscopes, Interoperable with Existing Equipment

Instruction:

Bidders must state in the column provided either “Comply” or “Not Comply” against EACH of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.

Item	Specification	Statement of Compliance
1	Nephroscope, Fr. 24 (with case and cleaning materials: brush and oil) 2 sets	
	1.1 Type:	
	1.1.1 Wide angle straight forward nephroscope	
	1.2 Scope Outer Diameter:	
	1.2.1 Outer diameter: 24 Fr.	
	1.3 Telescope/Viewing angle:	
	1.3.1 6° or 7° (six or thirty degrees view)	
	1.4 Eyepiece:	
	1.4.1 parallel eyepiece	
	1.5 Illumination:	
	1.5.1 Fiber optic light transmission incorporated with working channel	

	1.6 Operating Sheath 24 Fr.:	
	1.6.1 Operating sheath, 24 Fr., for continuous irrigation and suction, with luer lock stopcock, rotatable	
	1.7 Operating Sheath 26 Fr.:	
	1.7.1 Operating sheath, 26 Fr., for continuous irrigation and suction, with luer lock stopcock, rotatable	
	1.8 Obturator:	
	1.8.1 Hollow type obturator	
	1.9 Sealing Cap:	
	1.9.1 At least 5 pieces	
	1.10 Fascial Dilator:	
	1.10.1 Fascial Dilator	
	1.11 Processing	
	1.11.1 Autoclavable, Reprocessable	
	1.12 Compatibility	
	1.12.1 Compatible with 4 Fr. working instruments	
	1.13 Accessories per instrument (OEM):	
	1.13.1 Telescope bougie set, for tract dilation, consisting of: a) telescope bougie set, including 6 dilation sleeves 9, 12, 15, 18, 21 and 24 Fr., b) with 2 rigid guide wires and 2 flexible guide wires	
	1.13.2 Plastic container, for sterilizing and storage, perforated, with transparent lid, ,with inserts, external dimensions at least (w x d x h) 515 x 240 x 84	
	1.13.3 Cleaning brush and lubricant at least 10 pieces	
	1.13.4 Forceps, for grasping stone fragments, with fenestrated jaws and U-spring handle, 11.5 Fr., length 38cm at least one (1) piece	
	1.13.5 Forceps, for grasping larger stones and stone fragments, with triple serrated jaws and U-spring handle, 10.5 Fr., length 38cm at least one (1) piece	
	1.13.6 Forceps, for grasping larger stones and stone fragments, with serrated jaws and ring handle, double action jaws, 10.5 Fr., length 38cm at least one (1) piece	

	1.13.7 Forceps for grasping larger stones and stone fragments, with fenestrated jaws and ring handle, double action jaws, 10.5 Fr, length 38cm at least one (1) piece	
	1.13.8 Single action jaws Biopsy Forceps, with ring handle, 10.5 Fr., length 38cm at least one (1) piece	
2	Ureteroscope, Maximum of Fr. 7.8 (with case and cleaning materials: brush and oil), five (5) sets	
	2.1 Type:	
	1.1.1 Semi rigid, Fiber optic type	
	2.2 Working Channel:	
	2.2.1 Single, between 4.2 to 5 Fr., Straight channel	
	2.3 Telescope/Viewing angle:	
	2.3.1 between 6° to 7° (six to 7 degrees view)	
	2.4 Eyepiece:	
	2.4.1 Angled	
	2.5 Length:	
	2.5.1 Between 40 cm to 50 cm	
	2.6 Ports:	
	2.6.1 2 (two) lateral irrigation ports	
	2.7 Distal Tip:	
	2.7.1 between 6.5 to 7 Fr.	
	2.8 Processing:	
	2.8.1 Autoclavable, Reprocessable	
	2.9 Compatibility:	
	2.9.1 Compatible with 4 Fr. working instruments	
	2.10 Illumination	
	2.10.1 Fiber optic light transmission incorporated	

	2.11 Accessories per instrument (OEM):	
	2.11.1 Metal/Wire case for scope	
	2.11.2 Working channel	
	2.11.3 Adapters for working channel	
	2.11.4 Cleaning brush and lubricant at least 10 pieces	

Conforme:

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____